

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040323\
 Data File : PL081830.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Apr 2023 11:59
 Operator : AR\AJ
 Sample : 02126-04DL 5X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 RS-DDL

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 04/04/2023
 Supervised By : Ankita Jodhani 04/04/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 03 21:35:23 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 23 01:07:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.339	2.656	6948499	6253095	2.987	3.133
28)	SA Decachlor...	8.764	7.778	5779592	6214985	3.229m	3.134
Target Compounds							
10)	B gamma-Chl...	5.715	4.848	5981451	2959323	2.278m	1.218m#
11)	B alpha-Chl...	5.797	4.908	10735997	5593276	4.148	2.303m#
12)	B 4,4'-DDE	5.973	5.101	17257525	22269551	7.658	11.306 #
16)	A 4,4'-DDD	6.486	5.656	5789734	5040692	3.210m	3.322
17)	MA 4,4'-DDT	6.798	5.904	51252420	49358746	25.252	28.642m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040323\
Data File : PL081830.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Apr 2023 11:59
Operator : AR\AJ
Sample : 02126-04DL 5X
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
RS-DDL

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 04/04/2023
Supervised By : Ankita Jodhani 04/04/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 03 21:35:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
Quant Title : GC Extractables
QLast Update : Thu Mar 23 01:07:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

