

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040323\
 Data File : PL081832.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Apr 2023 12:50
 Operator : AR\AJ
 Sample : PB151853BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB151853BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 03 21:37:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 23 01:07:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.330	2.657	54849817	45696604	23.582	22.892
2)	SA Decachlor...	8.757	7.776	44218810	49358511	24.708	24.888
Target Compounds							
2)	A alpha-BHC	3.780	3.156	175.2E6	141.0E6	51.115	48.363
3)	MA gamma-BHC...	4.109	3.484	160.9E6	133.8E6	50.663	48.046
4)	MA Heptachlor	4.689	3.817	150.1E6	134.2E6	47.603	46.825
5)	MB Aldrin	5.026	4.095	149.0E6	134.4E6	52.104	50.431
6)	B beta-BHC	4.312	3.783	69107799	55411970	45.916	46.526
7)	B delta-BHC	4.553	4.011	160.5E6	132.1E6	52.823	51.097
8)	B Heptachlo...	5.451	4.597	136.0E6	124.5E6	50.009	51.253
9)	A Endosulfan I	5.833	4.965	120.3E6	115.0E6	48.702	52.107
10)	B gamma-Chl...	5.706	4.845	134.6E6	127.2E6	51.260	52.318
11)	B alpha-Chl...	5.784	4.909	131.0E6	125.7E6	50.600	51.742
12)	B 4,4'-DDE	5.963	5.100	117.1E6	107.4E6	51.942	54.517
13)	MA Dieldrin	6.107	5.229	127.7E6	121.5E6	52.733	54.007
14)	MA Endrin	6.335	5.503	106.9E6	105.2E6	50.526	52.008
15)	B Endosulfa...	6.557	5.800	107.4E6	104.4E6	50.595	57.537
16)	A 4,4'-DDD	6.479	5.655	95379987	91106154	52.886	60.036
17)	MA 4,4'-DDT	6.790	5.903	88939038	85838854	43.820	49.812
18)	B Endrin al...	6.687	5.980	89194742	86986483	54.516	56.012
19)	B Endosulfa...	6.921	6.203	105.2E6	104.9E6	50.462	53.918
20)	A Methoxychlor	7.271	6.483	49681055	49588087	46.134	49.785
21)	B Endrin ke...	7.401	6.708	108.3E6	115.2E6	51.709	53.889
22)	Mirex	7.866	6.880	84226085	99131219	49.161	51.638

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040323\
Data File : PL081832.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Apr 2023 12:50
Operator : AR\AJ
Sample : PB151853BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB151853BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 03 21:37:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
Quant Title : GC Extractables
QLast Update : Thu Mar 23 01:07:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

