

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040419\
 Data File : PL047030.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Apr 2019 16:04
 Operator : AJ\SJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 05 00:55:53 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.357	3.976	158.0E6	50959903	21.718	22.257
28)	SA Decachlor...	8.034	9.029	186.0E6	56760291	22.119	24.482
Target Compounds							
8)	B Heptachlo...	0.000	5.842	0	2633850	N.D.	0.999 #
10)	B gamma-Chl...	0.000	6.088	0	4809548	N.D.	1.695 #
12)	B 4,4'-DDE	5.504f	0.000	2741029	0	0.298	N.D. #
14)	MA Endrin	0.000	6.676	0	3244143	N.D.	1.358 #
25)	Chlordane-3	0.000	6.088	0	4809548	N.D.	12.615 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040419\
Data File : PL047030.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Apr 2019 16:04
Operator : AJ\SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 05 00:55:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.357 min
Delta R.T.: -0.004 min
Response: 157989189
Conc: 21.72 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.976 min
Delta R.T.: -0.004 min
Response: 50959903
Conc: 22.26 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 5.034 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.842 min
Delta R.T.: -0.012 min
Response: 2633850
Conc: 1.00 ng/ml



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T.: 5.257 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#10 gamma-Chlordane

R.T.: 6.088 min
Delta R.T.: 0.000 min
Response: 4809548
Conc: 1.69 ng/ml




#12 4,4'-DDE

R.T.: 5.504 min
Delta R.T.: 0.023 min
Response: 2741029
Conc: 0.30 ng/ml



#12 4,4'-DDE


R.T.: 0.000 min
Exp R.T.: 6.318 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 0.000 min
Exp R.T.: 5.860 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#14 Endrin

R.T.: 6.676 min
Delta R.T.: -0.008 min
Response: 3244143
Conc: 1.36 ng/ml




#25 Chlordane-3

R.T.: 0.000 min
Exp R.T.: 5.258 min
Response: 0
Conc: N.D.



#25 Chlordane-3

R.T.: 6.088 min
Delta R.T.: 0.000 min
Response: 4809548
Conc: 12.61 ng/ml



#28 Decachlorobiphenyl


R.T.: 8.034 min

Delta R.T.: -0.010 min

Response: 186040630

Conc: 22.12 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.029 min

Delta R.T.: -0.011 min

Response: 56760291

Conc: 24.48 ng/ml