

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040521\
 Data File : PL065862.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Apr 2021 10:07
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 05 14:48:52 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL033121.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 31 15:57:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.281	3.782	136.4E6	142.7E6	52.364	51.972
28)	SA Decachlor...	7.914	8.730	121.6E6	114.4E6	52.496	50.741
Target Compounds							
2)	A alpha-BHC	3.704	4.167	217.6E6	214.2E6	53.199	52.721
3)	MA gamma-BHC...	3.977	4.446	196.0E6	195.5E6	52.944	53.404
4)	MA Heptachlor	4.261	4.931	172.3E6	176.1E6	47.836	49.901
5)	MB Aldrin	4.498	5.228	179.3E6	174.8E6	52.517	52.934
6)	B beta-BHC	4.228	4.630	81130177	90408619	50.901	51.905
7)	B delta-BHC	4.420	4.837	193.1E6	169.0E6	53.671	56.655
8)	B Heptachlo...	4.930	5.614	163.3E6	159.2E6	51.930	51.276
9)	A Endosulfan I	5.257	5.964	151.7E6	144.8E6	52.649	52.758
10)	B gamma-Chl...	5.151	5.845	164.6E6	158.4E6	52.296	52.799
11)	B alpha-Chl...	5.206	5.914	161.1E6	156.3E6	51.788	52.804
12)	B 4,4'-DDE	5.373	6.079	146.9E6	141.8E6	53.243	54.277
13)	MA Dieldrin	5.494	6.221	154.1E6	149.5E6	52.728	53.052
14)	MA Endrin	5.744	6.436	124.1E6	116.9E6	45.975	46.937
15)	B Endosulfa...	6.016	6.653	128.2E6	123.7E6	51.626	52.323
16)	A 4,4'-DDD	5.879	6.571	129.0E6	122.1E6	56.861	57.948
17)	MA 4,4'-DDT	6.112	6.863	93412855	98887043	44.338	44.164
18)	B Endrin al...	6.182	6.777	97649846	99116716	51.609	52.236
19)	B Endosulfa...	6.392	6.998	132.6E6	124.7E6	51.390	51.401
20)	A Methoxychlor	6.657	7.328	52880944	52404199	42.124	41.376
21)	B Endrin ke...	6.869	7.464	146.1E6	139.2E6	53.278	54.157
22)	Mirex	7.041	7.899	106.2E6	97096998	50.692	49.747

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040521\
Data File : PL065862.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Apr 2021 10:07
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 05 14:48:52 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL033121.M
Quant Title : GC Extractables
QLast Update : Wed Mar 31 15:57:28 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

