

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040621\
 Data File : PL065903.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Apr 2021 13:04
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 4/7/2021 4:50:19 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 04:31:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL033121.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 31 15:57:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.276	3.780	133.6E6	137.0E6	51.289	49.927
28)	SA Decachlor...	7.907	8.727	124.5E6	118.5E6	53.755	52.550
Target Compounds							
2)	A alpha-BHC	3.699	4.165	212.9E6	204.9E6	52.061	50.441
3)	MA gamma-BHC...	3.970	4.444	192.9E6	187.1E6	52.100m	51.107
4)	MA Heptachlor	4.256	4.929	176.2E6	171.6E6	48.903	48.629
5)	MB Aldrin	4.491	5.225	176.6E6	168.1E6	51.718m	50.903
6)	B beta-BHC	4.222	4.627	81627093	86857961	51.212	49.867
7)	B delta-BHC	4.414	4.834	190.0E6	156.1E6	52.814	52.351
8)	B Heptachlo...	4.923	5.612	160.7E6	152.4E6	51.093	49.090
9)	A Endosulfan I	5.250	5.962	149.8E6	141.0E6	51.983	51.343
10)	B gamma-Chl...	5.145	5.843	161.8E6	153.2E6	51.412	51.072
11)	B alpha-Chl...	5.200	5.912	159.3E6	151.3E6	51.205	51.144
12)	B 4,4'-DDE	5.367	6.077	143.7E6	135.5E6	52.068	51.868
13)	MA Dieldrin	5.487	6.218	152.5E6	144.7E6	52.197	51.345
14)	MA Endrin	5.737	6.433	136.4E6	125.5E6	50.509	50.402
15)	B Endosulfa...	6.010	6.650	128.6E6	122.1E6	51.763	51.660
16)	A 4,4'-DDD	5.872	6.568	118.8E6	109.3E6	52.360	51.842
17)	MA 4,4'-DDT	6.104	6.860	102.7E6	102.2E6	48.764m	45.653
18)	B Endrin al...	6.176	6.774	97801992	96198104	51.689	50.698
19)	B Endosulfa...	6.385	6.996	132.8E6	122.5E6	51.468	50.513
20)	A Methoxychlor	6.649	7.327	59564689	56706685	47.448	44.773
21)	B Endrin ke...	6.863	7.462	141.1E6	130.2E6	51.443	50.645
22)	Mirex	7.035	7.897	107.1E6	97423951	51.141	49.914

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040621\
 Data File : PL065903.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Apr 2021 13:04
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 4/7/2021 4:50:19 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 04:31:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL033121.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 31 15:57:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

