

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040919\
 Data File : PL047121.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Apr 2019 16:05
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 01:02:49 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.358	3.977	373.5E6	119.6E6	51.340	52.223
28)	SA Decachlor...	8.036	9.032	368.7E6	109.6E6	43.833	47.276
Target Compounds							
2)	A alpha-BHC	3.787	4.367	621.3E6	171.6E6	53.280	53.413
3)	MA gamma-BHC...	4.065	4.652	556.6E6	159.0E6	51.805	53.137
4)	MA Heptachlor	4.351	5.158	556.5E6	158.7E6	50.473	51.592
5)	MB Aldrin	4.592	5.462	524.7E6	149.6E6	50.866	52.228
6)	B beta-BHC	4.317	4.825	223.9E6	70411310	49.724	53.113
7)	B delta-BHC	4.513	5.040	523.3E6	137.8E6	51.695	59.388
8)	B Heptachlor...	5.029	5.849	485.5E6	137.5E6	49.779	52.154
9)	A Endosulfan I	5.361	6.205	457.4E6	131.4E6	48.650	50.339
10)	B gamma-Chl...	5.252	6.083	500.5E6	144.8E6	48.722	51.033
11)	B alpha-Chl...	5.308	6.155	490.5E6	143.6E6	48.374	50.619
12)	B 4,4'-DDE	5.475	6.313	463.2E6	127.6E6	50.307	51.387
13)	MA Dieldrin	5.601	6.462	499.5E6	136.5E6	49.775	50.349
14)	MA Endrin	5.854	6.680	437.0E6	111.2E6	46.391	46.540
15)	B Endosulfa...	6.129	6.890	428.9E6	120.0E6	47.875	49.742
16)	A 4,4'-DDD	5.987	6.804	391.6E6	106.1E6	51.799	52.757
17)	MA 4,4'-DDT	6.222	7.101	351.1E6	98178289	46.348	46.819
18)	B Endrin al...	6.295	7.014	349.6E6	102.6E6	46.638	48.293
19)	B Endosulfa...	6.507	7.237	387.7E6	113.8E6	47.215	48.260
20)	A Methoxychlor	6.768	7.561	181.9E6	57870246	42.796	45.396
21)	B Endrin ke...	6.991	7.707	441.6E6	123.5E6	47.760	49.727
22)	Mirex	7.166	8.158	314.9E6	95673806	43.224	47.833

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040919\
Data File : PL047121.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Apr 2019 16:05
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 01:02:49 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

