

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040919\
 Data File : PL047158.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2019 01:12
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 01:24:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.358	3.977	347.3E6	112.2E6	47.737	48.993
28)	SA Decachlor...	8.037	9.032	376.4E6	108.2E6	44.750	46.672
Target Compounds							
2)	A alpha-BHC	3.788	4.368	584.2E6	162.8E6	50.095	50.675
3)	MA gamma-BHC...	4.065	4.652	521.5E6	151.2E6	48.532	50.527
4)	MA Heptachlor	4.351	5.158	528.6E6	151.0E6	47.943	49.103
5)	MB Aldrin	4.592	5.462	501.0E6	144.1E6	48.573	50.292
6)	B beta-BHC	4.318	4.825	213.3E6	67450051	47.383	50.879
7)	B delta-BHC	4.513	5.041	499.7E6	130.3E6	49.361	56.174
8)	B Heptachlor...	5.029	5.848	469.4E6	132.3E6	48.123	50.173
9)	A Endosulfan I	5.361	6.204	444.7E6	127.3E6	47.298	48.795
10)	B gamma-Chl...	5.252	6.082	487.4E6	139.6E6	47.446	49.188
11)	B alpha-Chl...	5.308	6.155	477.0E6	138.7E6	47.042	48.880
12)	B 4,4'-DDE	5.475	6.313	448.3E6	122.6E6	48.689	49.393
13)	MA Dieldrin	5.600	6.461	477.0E6	131.3E6	47.531	48.441
14)	MA Endrin	5.854	6.679	422.9E6	110.0E6	44.897	46.027
15)	B Endosulfa...	6.128	6.890	417.6E6	117.7E6	46.608	48.774
16)	A 4,4'-DDD	5.986	6.803	377.0E6	103.8E6	49.869	51.589
17)	MA 4,4'-DDT	6.221	7.101	342.0E6	95873881	45.150	45.721
18)	B Endrin al...	6.295	7.013	344.1E6	101.6E6	45.913	47.838
19)	B Endosulfa...	6.506	7.236	380.5E6	110.7E6	46.334	46.937
20)	A Methoxychlor	6.766	7.559	182.4E6	56686153	42.923	44.468
21)	B Endrin ke...	6.990	7.705	438.0E6	121.7E6	47.371	48.978
22)	Mirex	7.167	8.157	320.5E6	94807462	43.997	47.400

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040919\
Data File : PL047158.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2019 01:12
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 01:24:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

