

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040919\
 Data File : PL047165.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2019 03:07
 Operator : AJ\SJ
 Sample : K2317-14MS
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 COMP-1MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 04:02:52 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.358	3.977	148.9E6	50527553	20.463	22.068
28)	SA Decachlor...	8.036	9.034	167.8E6	49324156	19.952	21.274
Target Compounds							
2)	A alpha-BHC	3.788	4.367	660.7E6	188.7E6	56.655	58.728
3)	MA gamma-BHC...	4.065	4.652	605.2E6	177.0E6	56.324	59.137
4)	MA Heptachlor	4.352	5.158	609.4E6	176.5E6	55.277	57.388
5)	MB Aldrin	4.592	5.462	583.7E6	169.4E6	56.584	59.116
6)	B beta-BHC	4.318	4.825	243.6E6	78344996	54.105	59.097
7)	B delta-BHC	4.513	5.041	555.6E6	116.6E6	54.885	50.256
8)	B Heptachlor...	5.030	5.848	537.1E6	132.7E6	55.063	50.322
9)	A Endosulfan I	5.361	6.205	506.3E6	148.7E6	53.845	56.970
10)	B gamma-Chl...	5.253	6.082	557.4E6	160.4E6	54.263	56.516
11)	B alpha-Chl...	5.308	6.155	546.4E6	161.0E6	53.891	56.761
12)	B 4,4'-DDE	5.475	6.313	510.0E6	144.5E6	55.394	58.198
13)	MA Dieldrin	5.601	6.462	549.3E6	154.7E6	54.740	57.074
14)	MA Endrin	5.854	6.679	499.0E6	130.1E6	52.978	54.451
15)	B Endosulfa...	6.129	6.890	470.7E6	135.3E6	52.541	56.073
16)	A 4,4'-DDD	5.987	6.803	424.8E6	120.9E6	56.200	60.093
17)	MA 4,4'-DDT	6.222	7.102	408.0E6	116.2E6	53.870	55.404
18)	B Endrin al...	6.295	7.014	384.2E6	115.1E6	51.256	54.178
19)	B Endosulfa...	6.507	7.237	453.6E6	133.5E6	55.240	56.612
20)	A Methoxychlor	6.767	7.561	208.4E6	67943753	49.040	53.299
21)	B Endrin ke...	6.990	7.706	505.1E6	143.0E6	54.624	57.564
22)	Mirex	7.167	8.159	363.4E6	109.1E6	49.885	54.530

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040919\
Data File : PL047165.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2019 03:07
Operator : AJ\SJ
Sample : K2317-14MS
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
COMP-1MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 04:02:52 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

