

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040921\
 Data File : PL065980.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Apr 2021 23:11
 Operator : AR\AJ
 Sample : PTOXICV500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Integration File signal 1: events.e
 Integration File signal 2: events2.e
 Quant Time: Apr 12 01:13:53 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX040921.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 10 02:35:18 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. :
 Signal #1 Phase : Signal #2 Phase:
 Signal #1 Info : Signal #2 Info :

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA	Tetrachloro-m-xylene	2.347	2.274 E6	3.1	95	0.00
2	Toxaphene-1	19.406	20.178 E3	-4.0	101	0.00
3	Toxaphene-2	19.389	21.197 E3	-9.3	104	0.00
4	Toxaphene-3	62.306	60.652 E3	2.7	94	0.00
5	Toxaphene-4	42.372	40.159 E3	5.2	94	0.00
6	Toxaphene-5	53.165	49.501 E3	6.9	91	0.00
7 SA	Decachlorobiphenyl	2.111	2.017 E6	4.5	95	0.00

Signal #2

1 SA	Tetrachloro-m-xylene	2.667	2.646 E6	0.8	98	0.00
2	Toxaphene-1	11.411	12.311 E3	-7.9	105	0.00
3	Toxaphene-2	19.657	18.878 E3	4.0	95	0.00
4	Toxaphene-3	35.069	34.540 E3	1.5	99	0.00
5	Toxaphene-4	33.976	32.973 E3	3.0	97	0.00
6	Toxaphene-5	35.334	34.418 E3	2.6	97	0.00
7 SA	Decachlorobiphenyl	1.927	1.927 E6	0.0	99	0.00

Evaluate Continuing Calibration Report - Not Founds

Signal #2

(#) = Out of Range

SPCC's out = 0 CCC's out = 0