

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041019\
 Data File : PL047180.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2019 10:11
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 01:00:49 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.358	3.976	363.7E6	112.5E6	49.989	49.137
28)	SA Decachlor...	8.038	9.032	401.3E6	112.4E6	47.717	48.459
Target Compounds							
2)	A alpha-BHC	3.788	4.366	592.6E6	163.4E6	50.817	50.853
3)	MA gamma-BHC...	4.065	4.650	543.1E6	151.4E6	50.547	50.586
4)	MA Heptachlor	4.352	5.157	551.4E6	152.1E6	50.009	49.434
5)	MB Aldrin	4.592	5.461	526.5E6	144.5E6	51.045	50.444
6)	B beta-BHC	4.318	4.824	223.1E6	67605871	49.560	50.997
7)	B delta-BHC	4.514	5.039	526.1E6	132.1E6	51.968	56.914
8)	B Heptachlor...	5.030	5.847	491.1E6	136.5E6	50.346	51.769
9)	A Endosulfan I	5.361	6.203	465.0E6	128.3E6	49.461	49.160
10)	B gamma-Chl...	5.253	6.081	512.1E6	140.4E6	49.853	49.487
11)	B alpha-Chl...	5.308	6.154	499.7E6	139.6E6	49.288	49.209
12)	B 4,4'-DDE	5.476	6.312	476.2E6	123.5E6	51.727	49.747
13)	MA Dieldrin	5.601	6.460	500.8E6	132.8E6	49.903	48.986
14)	MA Endrin	5.854	6.678	435.7E6	110.3E6	46.257	46.175
15)	B Endosulfa...	6.129	6.888	438.7E6	119.1E6	48.960	49.361
16)	A 4,4'-DDD	5.988	6.802	398.5E6	104.7E6	52.725	52.052
17)	MA 4,4'-DDT	6.223	7.100	352.3E6	95256528	46.518	45.426
18)	B Endrin al...	6.296	7.012	363.1E6	103.0E6	48.443	48.491
19)	B Endosulfa...	6.508	7.235	402.1E6	112.7E6	48.968	47.784
20)	A Methoxychlor	6.767	7.559	188.8E6	56587968	44.418	44.391
21)	B Endrin ke...	6.990	7.705	465.2E6	124.3E6	50.312	50.013
22)	Mirex	7.167	8.157	341.1E6	96862195	46.825	48.427

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041019\
Data File : PL047180.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2019 10:11
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 01:00:49 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

