

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041019\
 Data File : PL047259.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Apr 2019 11:49
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 4/11/2019 3:57:32 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 12:01:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.355	3.974	376.0E6	114.4E6	51.683	49.960
28)	SA Decachlor...	8.029	9.024	422.2E6	122.3E6	50.200	52.740
Target Compounds							
2)	A alpha-BHC	3.784	4.364	614.2E6	164.6E6	52.669	51.219
3)	MA gamma-BHC...	4.059	4.648	534.3E6	153.5E6	49.725m	51.299
4)	MA Heptachlor	4.348	5.153	583.6E6	160.4E6	52.931	52.148m
5)	MB Aldrin	4.588	5.458	538.3E6	146.5E6	52.186	51.142
6)	B beta-BHC	4.314	4.821	222.8E6	67289049	49.495	50.758
7)	B delta-BHC	4.509	5.036	539.6E6	114.4E6	53.300	49.298m
8)	B Heptachlor...	5.024	5.842	504.0E6	120.2E6	51.676	45.560m
9)	A Endosulfan I	5.356	6.200	483.8E6	131.4E6	51.451	50.338
10)	B gamma-Chl...	5.247	6.078	529.4E6	140.3E6	51.537	49.425
11)	B alpha-Chl...	5.304	6.150	519.5E6	142.5E6	51.234	50.252
12)	B 4,4'-DDE	5.469	6.308	461.2E6	126.0E6	50.090m	50.762
13)	MA Dieldrin	5.596	6.455	518.9E6	136.2E6	51.705	50.229m
14)	MA Endrin	5.849	6.674	501.5E6	121.4E6	53.244	50.812
15)	B Endosulfa...	6.123	6.884	465.6E6	117.9E6	51.962	48.869
16)	A 4,4'-DDD	5.982	6.799	391.2E6	100.8E6	51.754	50.100
17)	MA 4,4'-DDT	6.217	7.095	413.2E6	110.0E6	54.552	52.447m
18)	B Endrin al...	6.290	7.008	387.0E6	104.8E6	51.640	49.331
19)	B Endosulfa...	6.502	7.230	431.8E6	115.3E6	52.581	48.868
20)	A Methoxychlor	6.760	7.555	226.9E6	66366199	53.380m	52.061
21)	B Endrin ke...	6.983	7.700	470.4E6	124.9E6	50.870m	50.287
22)	Mirex	7.161	8.151	348.8E6	97462368	47.870	48.727

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041019\
Data File : PL047259.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Apr 2019 11:49
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

mohammad
4/11/2019 3:57:32 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 12:01:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

