

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041023\
 Data File : PL081973.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2023 21:23
 Operator : AR\AJ
 Sample : 02224-07
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 OU3-FC-BACKFILL-04

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 23:06:21 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 23 01:07:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.325	2.653	41036962	37513871	17.644	18.793
28)	SA Decachlor...	8.744	7.767	32000353	35468574	17.881	17.884

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041023\
Data File : PL081973.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2023 21:23
Operator : AR\AJ
Sample : 02224-07
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
OU3-FC-BACKFILL-04

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 23:06:21 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
Quant Title : GC Extractables
QLast Update : Thu Mar 23 01:07:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.325 min
Delta R.T.: -0.010 min
Response: 41036962
Conc: 17.64 ng/ml

Instrument :
ECD_L
ClientSampleId :
OU3-FC-BACKFILL-04




#1 Tetrachloro-m-xylene

R.T.: 2.653 min
Delta R.T.: -0.003 min
Response: 37513871
Conc: 18.79 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.744 min
Delta R.T.: -0.016 min
Response: 32000353
Conc: 17.88 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.767 min
Delta R.T.: -0.012 min
Response: 35468574
Conc: 17.88 ng/ml