

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041024\
 Data File : PL088977.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Apr 2024 20:30
 Operator : AR\AJ
 Sample : INDB201
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 INDB2254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 12 04:59:17 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041024CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 12 04:53:41 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.634	2.851	26223200	14126725	10.079	9.992
27)	SA Decachlor...	9.299	8.077	51386889	33695524	20.820	20.966
Target Compounds							
5)	MB Aldrin	5.383	4.335	33701405	17790632	9.931	9.725
6)	B beta-BHC	4.634	4.003	17545449	9624064	10.196	10.288
7)	B delta-BHC	4.885	4.238	38267409	17873006	10.860	9.406
8)	B Heptachlo...	5.815	4.845	38662058	18341646	11.630	10.440
10)	B trans-Chl...	6.073	5.099	32754699	17439206	10.163	9.731
11)	B cis-Chlor...	6.154	5.164	33577648	18189072	10.184	9.979
12)	B 4,4'-DDE	6.329	5.357	58140564	31446478	19.489	18.888
15)	B Endosulfa...	6.937	6.066	59634759	31761320	20.845	19.917
18)	B Endrin al...	7.068	6.248	45772973	25746067	20.539	20.255
19)	B Endosulfa...	7.305	6.474	55932724	31305023	20.525	20.096
21)	B Endrin ke...	7.797	6.986	59023235	35758782	20.260	19.989


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041024\
Data File : PL088977.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Apr 2024 20:30
Operator : AR\AJ
Sample : INDB201
Misc :
ALS Vial : 12 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 12 04:59:17 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Apr 12 04:53:41 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.634 min
Delta R.T.: 0.000 min
Response: 26223200
Conc: 10.08 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDB2254




#1 Tetrachloro-m-xylene

R.T.: 2.851 min
Delta R.T.: 0.000 min
Response: 14126725
Conc: 9.99 ng/ml




#5 Aldrin

R.T.: 5.383 min
Delta R.T.: 0.000 min
Response: 33701405
Conc: 9.93 ng/ml



#5 Aldrin


R.T.: 4.335 min
Delta R.T.: 0.000 min
Response: 17790632
Conc: 9.73 ng/ml



#6 beta-BHC


R.T.: 4.634 min
Delta R.T.: 0.000 min
Response: 17545449
Conc: 10.20 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDB2254




#6 beta-BHC

R.T.: 4.003 min
Delta R.T.: 0.000 min
Response: 9624064
Conc: 10.29 ng/ml




#7 delta-BHC

R.T.: 4.885 min
Delta R.T.: 0.000 min
Response: 38267409
Conc: 10.86 ng/ml



#7 delta-BHC


R.T.: 4.238 min
Delta R.T.: 0.000 min
Response: 17873006
Conc: 9.41 ng/ml



#8 Heptachlor epoxide


R.T.: 5.815 min
Delta R.T.: 0.000 min
Response: 38662058
Conc: 11.63 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDB2254




#8 Heptachlor epoxide

R.T.: 4.845 min
Delta R.T.: 0.000 min
Response: 18341646
Conc: 10.44 ng/ml




#10 trans-Chlordane

R.T.: 6.073 min
Delta R.T.: 0.000 min
Response: 32754699
Conc: 10.16 ng/ml



#10 trans-Chlordane


R.T.: 5.099 min
Delta R.T.: 0.000 min
Response: 17439206
Conc: 9.73 ng/ml



#11 cis-Chlordane


R.T.: 6.154 min
Delta R.T.: 0.000 min
Response: 33577648
Conc: 10.18 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDB2254




#11 cis-Chlordane

R.T.: 5.164 min
Delta R.T.: 0.000 min
Response: 18189072
Conc: 9.98 ng/ml




#12 4,4'-DDE

R.T.: 6.329 min
Delta R.T.: 0.000 min
Response: 58140564
Conc: 19.49 ng/ml



#12 4,4'-DDE


R.T.: 5.357 min
Delta R.T.: 0.000 min
Response: 31446478
Conc: 18.89 ng/ml



#15 Endosulfan II


R.T.: 6.937 min
Delta R.T.: 0.000 min
Response: 59634759
Conc: 20.85 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDB2254




#15 Endosulfan II

R.T.: 6.066 min
Delta R.T.: 0.000 min
Response: 31761320
Conc: 19.92 ng/ml




#18 Endrin aldehyde

R.T.: 7.068 min
Delta R.T.: -0.001 min
Response: 45772973
Conc: 20.54 ng/ml



#18 Endrin aldehyde


R.T.: 6.248 min
Delta R.T.: 0.000 min
Response: 25746067
Conc: 20.26 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.305 min
Delta R.T.: 0.000 min
Response: 55932724
Conc: 20.52 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDB2254




#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 31305023
Conc: 20.10 ng/ml




#21 Endrin ketone

R.T.: 7.797 min
Delta R.T.: 0.000 min
Response: 59023235
Conc: 20.26 ng/ml



#21 Endrin ketone


R.T.: 6.986 min
Delta R.T.: 0.000 min
Response: 35758782
Conc: 19.99 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.299 min
Delta R.T.: 0.000 min
Response: 51386889
Conc: 20.82 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDB2254



#27 Decachlorobiphenyl

R.T.: 8.077 min
Delta R.T.: 0.000 min
Response: 33695524
Conc: 20.97 ng/ml