

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041119\
 Data File : PL047260.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Apr 2019 12:09
 Operator : AJ\SJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 12 01:08:06 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds

6) B	beta-BHC	0.000	4.847f	0	3428961	N.D.	2.587 #
18) B	Endrin al...	6.298	0.000	7262948	0	0.969	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041119\
Data File : PL047260.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Apr 2019 12:09
Operator : AJ\SJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 12 01:08:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 3.361 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
HEXANE




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.981 min
Response: 0
Conc: N.D.




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.321 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 4.847 min
Delta R.T.: 0.019 min
Response: 3428961
Conc: 2.59 ng/ml



#18 Endrin aldehyde


R.T.: 6.298 min
Delta R.T.: -0.004 min
Response: 7262948
Conc: 0.97 ng/ml

Instrument :
ECD_L
ClientSampleId :
HEXANE




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T. : 7.019 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.044 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 9.035 min
Response: 0
Conc: N.D.