

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041219\
 Data File : PL047363.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Apr 2019 00:52
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 4/15/2019 9:28:06 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 13 05:42:44 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.356	3.975	401.7E6	123.5E6	55.222	53.961
28)	SA Decachlor...	8.031	9.025	423.1E6	123.5E6	50.306	53.259
Target Compounds							
2)	A alpha-BHC	3.786	4.364	680.8E6	177.5E6	58.377	55.232
3)	MA gamma-BHC...	4.063	4.648	607.0E6	165.3E6	56.497	55.220
4)	MA Heptachlor	4.350	5.153	622.2E6	169.5E6	56.438	55.118m
5)	MB Aldrin	4.589	5.458	580.1E6	154.9E6	56.237	54.074
6)	B beta-BHC	4.315	4.822	245.3E6	72304194	54.471	54.541
7)	B delta-BHC	4.511	5.036	578.5E6	137.7E6	57.148	59.358m
8)	B Heptachlor...	5.026	5.843	533.8E6	138.7E6	54.726	52.590m
9)	A Endosulfan I	5.358	6.201	501.6E6	136.5E6	53.345	52.313
10)	B gamma-Chl...	5.249	6.079	554.2E6	148.8E6	53.947	52.442
11)	B alpha-Chl...	5.305	6.151	541.7E6	148.6E6	53.425	52.389
12)	B 4,4'-DDE	5.471	6.309	441.7E6	130.5E6	47.970m	52.567
13)	MA Dieldrin	5.597	6.456	534.5E6	141.1E6	53.262	52.071m
14)	MA Endrin	5.850	6.675	498.4E6	125.2E6	52.910	52.403
15)	B Endosulfa...	6.125	6.885	472.5E6	125.6E6	52.732	52.062
16)	A 4,4'-DDD	5.983	6.799	414.3E6	106.0E6	54.811	52.706
17)	MA 4,4'-DDT	6.218	7.095	410.5E6	109.7E6	54.192	52.320m
18)	B Endrin al...	6.291	7.009	383.1E6	107.7E6	51.107	50.725
19)	B Endosulfa...	6.503	7.232	423.8E6	119.0E6	51.609	50.432
20)	A Methoxychlor	6.762	7.555	214.3E6	64841662	50.424	50.865
21)	B Endrin ke...	6.984	7.701	481.2E6	128.5E6	52.041m	51.703
22)	Mirex	7.161	8.153	353.4E6	101.8E6	48.509	50.912

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041219\
Data File : PL047363.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Apr 2019 00:52
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
4/15/2019 9:28:06 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 13 05:42:44 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

