

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041224\
 Data File : PL089006.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2024 15:48
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 16:00:27 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031324.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 13 15:45:55 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.634	2.851	127.8E6	68994225	48.239	51.769
28)	SA Decachlor...	9.301	8.078	118.8E6	77683799	48.045	57.292
Target Compounds							
2)	A alpha-BHC	4.099	3.366	197.5E6	106.8E6	49.468	53.143
3)	MA gamma-BHC...	4.437	3.702	191.3E6	103.4E6	49.315	53.904
4)	MA Heptachlor	5.037	4.050	186.8E6	99924033	49.201	55.522
5)	MB Aldrin	5.385	4.336	181.2E6	100.8E6	49.446	55.767
6)	B beta-BHC	4.635	4.004	82496811	44119045	46.518	51.466
7)	B delta-BHC	4.887	4.239	187.3E6	101.2E6	46.506	52.885
8)	B Heptachlo...	5.816	4.846	174.1E6	90252436	50.692	54.881
9)	A Endosulfan I	6.207	5.222	151.4E6	81551807	49.520	54.250
10)	B gamma-Chl...	6.075	5.101	170.1E6	91339490	51.083	55.146
11)	B alpha-Chl...	6.156	5.165	166.8E6	91640482	50.191	55.207
12)	B 4,4'-DDE	6.330	5.358	147.2E6	82724076	49.271	56.715
13)	MA Dieldrin	6.486	5.491	160.4E6	90427241	49.466	55.677
14)	MA Endrin	6.718	5.770	139.1E6	80676708	50.878	55.970
15)	B Endosulfa...	6.938	6.067	140.2E6	77993621	48.944	54.242
16)	A 4,4'-DDD	6.852	5.920	125.8E6	68823958	49.392	55.264
17)	MA 4,4'-DDT	7.171	6.175	131.3E6	76332906	49.248	56.872
18)	B Endrin al...	7.070	6.250	86101354	59511055	41.240	56.788 #
19)	B Endosulfa...	7.307	6.475	133.7E6	76036495	48.412	54.803
20)	A Methoxychlor	7.652	6.758	74310498	44436222	49.326	55.402
21)	B Endrin ke...	7.798	6.987	145.8E6	89102535	49.332	53.978
22)	Mirex	8.280	7.173	116.5E6	74066023	49.703	54.928
23)	Chlordane-1	0.000	3.882	0	145745	N.D.	3.002 #
24)	Chlordane-2	5.385f	4.474	181.2E6	482664	1378.224	7.950 #
25)	Chlordane-3	6.075	5.101	170.1E6	91339490	342.191	535.180 #
26)	Chlordane-4	6.156	5.165	166.8E6	91640482	278.122	574.405 #
27)	Chlordane-5	0.000	5.866f	0	164938	N.D.	3.252 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041224\
Data File : PL089006.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2024 15:48
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 16:00:27 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031324.M
Quant Title : GC Extractables
QLast Update : Wed Mar 13 15:45:55 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.634 min
Delta R.T.: 0.000 min
Response: 127831749
Conc: 48.24 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.851 min
Delta R.T.: 0.000 min
Response: 68994225
Conc: 51.77 ng/ml




#2 alpha-BHC

R.T.: 4.099 min
Delta R.T.: 0.000 min
Response: 197461046
Conc: 49.47 ng/ml



#2 alpha-BHC


R.T.: 3.366 min
Delta R.T.: -0.001 min
Response: 106788947
Conc: 53.14 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.437 min
Delta R.T.: 0.000 min
Response: 191255691
Conc: 49.32 ng/ml

Instrument :
ECD_L
ClientSampleId :




#3 gamma-BHC (Lindane)

R.T.: 3.702 min
Delta R.T.: -0.001 min
Response: 103448799
Conc: 53.90 ng/ml




#4 Heptachlor

R.T.: 5.037 min
Delta R.T.: 0.000 min
Response: 186838698
Conc: 49.20 ng/ml



#4 Heptachlor


R.T.: 4.050 min
Delta R.T.: -0.002 min
Response: 99924033
Conc: 55.52 ng/ml



#5 Aldrin


R.T.: 5.385 min
Delta R.T.: 0.000 min
Response: 181192850
Conc: 49.45 ng/ml

Instrument :
ECD_L
ClientSampleId :




#5 Aldrin

R.T.: 4.336 min
Delta R.T.: -0.002 min
Response: 100775412
Conc: 55.77 ng/ml




#6 beta-BHC

R.T.: 4.635 min
Delta R.T.: 0.000 min
Response: 82496811
Conc: 46.52 ng/ml



#6 beta-BHC


R.T.: 4.004 min
Delta R.T.: -0.002 min
Response: 44119045
Conc: 51.47 ng/ml



#7 delta-BHC


R.T.: 4.887 min
Delta R.T.: 0.000 min
Response: 187336266
Conc: 46.51 ng/ml

Instrument :
ECD_L
ClientSampleId :




#7 delta-BHC

R.T.: 4.239 min
Delta R.T.: -0.002 min
Response: 101172491
Conc: 52.88 ng/ml




#8 Heptachlor epoxide

R.T.: 5.816 min
Delta R.T.: -0.001 min
Response: 174125337
Conc: 50.69 ng/ml



#8 Heptachlor epoxide


R.T.: 4.846 min
Delta R.T.: -0.003 min
Response: 90252436
Conc: 54.88 ng/ml



#9 Endosulfan I


R.T.: 6.207 min
Delta R.T.: 0.000 min
Response: 151356844
Conc: 49.52 ng/ml

Instrument :
ECD_L
ClientSampleId :




#9 Endosulfan I

R.T.: 5.222 min
Delta R.T.: -0.002 min
Response: 81551807
Conc: 54.25 ng/ml




#10 gamma-Chlordane

R.T.: 6.075 min
Delta R.T.: 0.000 min
Response: 170149594
Conc: 51.08 ng/ml



#10 gamma-Chlordane


R.T.: 5.101 min
Delta R.T.: -0.003 min
Response: 91339490
Conc: 55.15 ng/ml



#11 alpha-Chlordane


R.T.: 6.156 min
Delta R.T.: -0.001 min
Response: 166836022
Conc: 50.19 ng/ml

Instrument :
ECD_L
ClientSampleId :




#11 alpha-Chlordane

R.T.: 5.165 min
Delta R.T.: -0.002 min
Response: 91640482
Conc: 55.21 ng/ml




#12 4,4'-DDE

R.T.: 6.330 min
Delta R.T.: -0.001 min
Response: 147207452
Conc: 49.27 ng/ml



#12 4,4'-DDE


R.T.: 5.358 min
Delta R.T.: -0.003 min
Response: 82724076
Conc: 56.72 ng/ml



#13 Dieldrin


R.T.: 6.486 min
Delta R.T.: -0.001 min
Response: 160402924
Conc: 49.47 ng/ml

Instrument :
ECD_L
ClientSampleId :




#13 Dieldrin

R.T.: 5.491 min
Delta R.T.: -0.003 min
Response: 90427241
Conc: 55.68 ng/ml




#14 Endrin

R.T.: 6.718 min
Delta R.T.: -0.002 min
Response: 139086004
Conc: 50.88 ng/ml



#14 Endrin


R.T.: 5.770 min
Delta R.T.: -0.003 min
Response: 80676708
Conc: 55.97 ng/ml



#15 Endosulfan II


R.T.: 6.938 min
Delta R.T.: -0.002 min
Response: 140185737
Conc: 48.94 ng/ml

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 6.067 min
Delta R.T.: -0.003 min
Response: 77993621
Conc: 54.24 ng/ml




#16 4,4'-DDD

R.T.: 6.852 min
Delta R.T.: -0.002 min
Response: 125789387
Conc: 49.39 ng/ml



#16 4,4'-DDD


R.T.: 5.920 min
Delta R.T.: -0.003 min
Response: 68823958
Conc: 55.26 ng/ml



#17 4,4'-DDT


R.T.: 7.171 min
Delta R.T.: -0.002 min
Response: 131270073
Conc: 49.25 ng/ml

Instrument :
ECD_L
ClientSampleId :




#17 4,4'-DDT

R.T.: 6.175 min
Delta R.T.: -0.003 min
Response: 76332906
Conc: 56.87 ng/ml




#18 Endrin aldehyde

R.T.: 7.070 min
Delta R.T.: -0.001 min
Response: 86101354
Conc: 41.24 ng/ml



#18 Endrin aldehyde


R.T.: 6.250 min
Delta R.T.: -0.003 min
Response: 59511055
Conc: 56.79 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.307 min
Delta R.T.: -0.002 min
Response: 133680980
Conc: 48.41 ng/ml

Instrument :
ECD_L
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 6.475 min
Delta R.T.: -0.003 min
Response: 76036495
Conc: 54.80 ng/ml




#20 Methoxychlor

R.T.: 7.652 min
Delta R.T.: -0.001 min
Response: 74310498
Conc: 49.33 ng/ml



#20 Methoxychlor


R.T.: 6.758 min
Delta R.T.: -0.003 min
Response: 44436222
Conc: 55.40 ng/ml



#21 Endrin ketone


R.T.: 7.798 min
Delta R.T.: -0.002 min
Response: 145846163
Conc: 49.33 ng/ml

Instrument :
ECD_L
ClientSampleId :




#21 Endrin ketone

R.T.: 6.987 min
Delta R.T.: -0.003 min
Response: 89102535
Conc: 53.98 ng/ml




#22 Mirex

R.T.: 8.280 min
Delta R.T.: -0.002 min
Response: 116527769
Conc: 49.70 ng/ml



#22 Mirex


R.T.: 7.173 min
Delta R.T.: -0.004 min
Response: 74066023
Conc: 54.93 ng/ml



#23 Chlordane-1


R.T.: 0.000 min
Exp R.T.: 4.819 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#23 Chlordane-1

R.T.: 3.882 min
Delta R.T.: 0.008 min
Response: 145745
Conc: 3.00 ng/ml




#24 Chlordane-2

R.T.: 5.385 min
Delta R.T.: 0.030 min
Response: 181192850
Conc: 1378.22 ng/ml



#24 Chlordane-2


R.T.: 4.474 min
Delta R.T.: 0.012 min
Response: 482664
Conc: 7.95 ng/ml



#25 Chlordane-3


R.T.: 6.075 min
Delta R.T.: -0.001 min
Response: 170149594
Conc: 342.19 ng/ml

Instrument :
ECD_L
ClientSampleId :




#25 Chlordane-3

R.T.: 5.101 min
Delta R.T.: -0.002 min
Response: 91339490
Conc: 535.18 ng/ml




#26 Chlordane-4

R.T.: 6.156 min
Delta R.T.: -0.004 min
Response: 166836022
Conc: 278.12 ng/ml



#26 Chlordane-4


R.T.: 5.165 min
Delta R.T.: -0.002 min
Response: 91640482
Conc: 574.41 ng/ml



#27 Chlordane-5


R.T.: 0.000 min
Exp R.T.: 7.019 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#27 Chlordane-5

R.T.: 5.866 min
Delta R.T.: 0.022 min
Response: 164938
Conc: 3.25 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.301 min
Delta R.T.: -0.002 min
Response: 118762462
Conc: 48.04 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.078 min
Delta R.T.: -0.004 min
Response: 77683799
Conc: 57.29 ng/ml