

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041420\
 Data File : PL057973.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Apr 2020 18:19
 Operator : AJ\MA
 Sample : L2115-08
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 15 01:01:12 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 04 06:41:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.457	3.956	203.4E6	43591133	20.180	20.087
28)	SA Decachlor...	8.203	8.978	150.5E6	34328409	14.596	15.368
Target Compounds							
4)	MA Heptachlor	0.000	5.143	0	2805276	N.D.	0.836 #
8)	B Heptachlo...	0.000	5.794f	0	8381782	N.D.	3.104 #
15)	B Endosulfa...	0.000	6.867	0	559596	N.D.	0.216 #
22)	Mirex	7.332	0.000	381816	0	0.047	N.D. #
23)	Chlordane-1	4.356f	4.936	2949663	4939158	7.975	56.010 #
27)	Chlordane-5	0.000	6.925	0	247754	N.D.	2.906 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041420\
Data File : PL057973.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2020 18:19
Operator : AJ\MA
Sample : L2115-08
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 15 01:01:12 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
Quant Title : GC Extractables
QLast Update : Sat Apr 04 06:41:11 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.457 min
Delta R.T.: 0.000 min
Response: 203435585
Conc: 20.18 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.956 min
Delta R.T.: -0.001 min
Response: 43591133
Conc: 20.09 ng/ml




#4 Heptachlor

R.T.: 0.000 min
Exp R.T. : 4.479 min
Response: 0
Conc: N.D.



#4 Heptachlor


R.T.: 5.143 min
Delta R.T.: 0.014 min
Response: 2805276
Conc: 0.84 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T. : 5.169 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.794 min
Delta R.T.: -0.021 min
Response: 8381782
Conc: 3.10 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 6.282 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 6.867 min
Delta R.T.: 0.015 min
Response: 559596
Conc: 0.22 ng/ml



#22 Mirex


R.T.: 7.332 min
Delta R.T.: -0.008 min
Response: 381816
Conc: 0.05 ng/ml

Instrument :
ECD_L
ClientSampleId :




#22 Mirex

R.T.: 0.000 min
Exp R.T. : 8.117 min
Response: 0
Conc: N.D.




#23 Chlordane-1

R.T.: 4.356 min
Delta R.T.: 0.029 min
Response: 2949663
Conc: 7.98 ng/ml



#23 Chlordane-1


R.T.: 4.936 min
Delta R.T.: -0.006 min
Response: 4939158
Conc: 56.01 ng/ml



#27 Chlordane-5


R.T.: 0.000 min
Exp R.T.: 6.290 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#27 Chlordane-5

R.T.: 6.925 min
Delta R.T.: 0.003 min
Response: 247754
Conc: 2.91 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.203 min
Delta R.T.: -0.003 min
Response: 150459112
Conc: 14.60 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.978 min
Delta R.T.: -0.003 min
Response: 34328409
Conc: 15.37 ng/ml