

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041323\
 Data File : PL082016.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Apr 2023 14:53
 Operator : AR\AJ
 Sample : TOXAPHENE301
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 13 09:10:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041323CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 13 09:07:43 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.325	2.652	53508520	50982678	20.000	20.000
27)	SA Decachlor...	8.746	7.767	77717068	100.4E6	40.000	40.000
Target Compounds							
22)	Toxaphene-1	5.996	5.187	32686411	22280373	2000.000	2000.000
23)	Toxaphene-2	6.812	5.544	123.0E6	24625881	2000.000	2000.000
24)	Toxaphene-3	6.898	6.584	87617601	142.1E6	2000.000	2000.000
25)	Toxaphene-4	6.996	6.895	53101702	89869910	2000.000	2000.000
26)	Toxaphene-5	7.679	7.027	65397164	46708502	2000.000	2000.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041323\
Data File : PL082016.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Apr 2023 14:53
Operator : AR\AJ
Sample : TOXAPHENE301
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 13 09:10:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041323CLP.M
Quant Title : GC Extractables
QLast Update : Thu Apr 13 09:07:43 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

