

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041420\
 Data File : PL057952.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Apr 2020 12:40
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 15 01:07:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 04 06:41:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.456	3.956	460.4E6	97873723	45.670	45.100
28)	SA Decachlor...	8.202	8.976	453.3E6	96514248	43.974	43.207
Target Compounds							
2)	A alpha-BHC	3.893	4.343	794.2E6	151.9E6	45.715	46.257
3)	MA gamma-BHC...	4.177	4.625	715.2E6	149.6E6	49.207	46.047
4)	MA Heptachlor	4.477	5.126	732.2E6	151.6E6	45.677	45.180
5)	MB Aldrin	4.724	5.428	694.5E6	147.4E6	47.165	45.510
6)	B beta-BHC	4.429	4.798	289.4E6	64840998	45.826	46.695
7)	B delta-BHC	4.632	5.012	720.8E6	106.1E6	47.497	42.753
8)	B Heptachlo...	5.167	5.812	624.5E6	123.1E6	46.464	45.571
9)	A Endosulfan I	5.505	6.167	563.5E6	134.2E6	45.971	45.001
10)	B gamma-Chl...	5.393	6.046	624.1E6	143.5E6	45.953	44.278
11)	B alpha-Chl...	5.451	6.118	604.5E6	147.1E6	45.395	44.966
12)	B 4,4'-DDE	5.615	6.277	598.7E6	133.8E6	45.703	45.176
13)	MA Dieldrin	5.748	6.422	603.8E6	138.3E6	46.415	45.245
14)	MA Endrin	6.005	6.639	533.4E6	109.6E6	45.698	43.871
15)	B Endosulfa...	6.278	6.849	502.5E6	115.8E6	45.424	44.665
16)	A 4,4'-DDD	6.131	6.765	492.7E6	107.7E6	49.297	47.174
17)	MA 4,4'-DDT	6.371	7.061	427.5E6	92785177	42.119	41.657
18)	B Endrin al...	6.447	6.973	409.8E6	96321126	43.832	43.356
19)	B Endosulfa...	6.660	7.195	459.4E6	103.6E6	45.218	43.677
20)	A Methoxychlor	6.916	7.521	231.6E6	51472557	42.020	40.465
21)	B Endrin ke...	7.150	7.663	521.2E6	112.5E6	46.814	45.508
22)	Mirex	7.336	8.112	355.5E6	80670759	43.372	43.876

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041420\
Data File : PL057952.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2020 12:40
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 15 01:07:59 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
Quant Title : GC Extractables
QLast Update : Sat Apr 04 06:41:11 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

