

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041420\
 Data File : PL057962.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Apr 2020 15:43
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 15 01:11:16 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 04 06:41:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.457	3.956	490.1E6	100.9E6	48.617	46.499
28)	SA Decachlor...	8.203	8.976	469.7E6	98622337	45.562	44.151
Target Compounds							
2)	A alpha-BHC	3.894	4.343	851.3E6	157.7E6	49.006	48.033
3)	MA gamma-BHC...	4.177	4.625	763.5E6	155.8E6	52.531	47.961
4)	MA Heptachlor	4.477	5.127	795.5E6	158.9E6	49.623	47.336
5)	MB Aldrin	4.725	5.428	754.6E6	154.8E6	51.245	47.791
6)	B beta-BHC	4.429	4.799	315.7E6	67734661	49.990	48.779
7)	B delta-BHC	4.633	5.013	785.2E6	119.6E6	51.736	47.920
8)	B Heptachlo...	5.168	5.812	682.7E6	138.0E6	50.796	51.118
9)	A Endosulfan I	5.506	6.167	620.3E6	140.6E6	50.613	47.148
10)	B gamma-Chl...	5.394	6.046	683.5E6	153.5E6	50.327	47.373
11)	B alpha-Chl...	5.452	6.118	667.5E6	154.6E6	50.130	47.240
12)	B 4,4'-DDE	5.616	6.277	650.1E6	140.4E6	49.626	47.429
13)	MA Dieldrin	5.749	6.423	657.9E6	144.9E6	50.572	47.413
14)	MA Endrin	6.005	6.639	564.5E6	114.5E6	48.359	45.834
15)	B Endosulfa...	6.279	6.850	541.9E6	118.4E6	48.984	45.657
16)	A 4,4'-DDD	6.131	6.766	521.8E6	112.0E6	52.210	49.069
17)	MA 4,4'-DDT	6.371	7.062	455.5E6	95152012	44.877	42.719
18)	B Endrin al...	6.447	6.973	439.9E6	99810249	47.054	44.926
19)	B Endosulfa...	6.660	7.196	495.5E6	108.2E6	48.769	45.577
20)	A Methoxychlor	6.917	7.522	233.5E6	52173637	42.364	41.016
21)	B Endrin ke...	7.151	7.664	554.6E6	116.0E6	49.815	46.922
22)	Mirex	7.338	8.113	382.0E6	83143188	46.605	45.220

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041420\
Data File : PL057962.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2020 15:43
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 15 01:11:16 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
Quant Title : GC Extractables
QLast Update : Sat Apr 04 06:41:11 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

