

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041420\
 Data File : PL058002.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Apr 2020 01:26
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 15 01:53:09 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 04 06:41:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.458	3.956	490.3E6	105.1E6	48.634	48.421
2)	SA Decachlor...	8.204	8.977	448.9E6	99935151	43.542	44.738
Target Compounds							
2)	A alpha-BHC	3.895	4.344	859.6E6	164.3E6	49.484	50.025
3)	MA gamma-BHC...	4.178	4.626	784.2E6	161.3E6	53.950	49.639
4)	MA Heptachlor	4.478	5.128	799.9E6	161.6E6	49.899	48.152
5)	MB Aldrin	4.726	5.429	763.4E6	159.1E6	51.842	49.118
6)	B beta-BHC	4.430	4.799	321.6E6	68915923	50.919	49.630
7)	B delta-BHC	4.633	5.013	793.7E6	132.7E6	52.301	52.933
8)	B Heptachlo...	5.168	5.814	692.9E6	145.9E6	51.558	54.023
9)	A Endosulfan I	5.506	6.168	625.7E6	142.3E6	51.051	47.721
10)	B gamma-Chl...	5.395	6.047	694.9E6	156.5E6	51.164	48.306
11)	B alpha-Chl...	5.453	6.119	676.1E6	156.3E6	50.773	47.777
12)	B 4,4'-DDE	5.617	6.278	652.0E6	141.9E6	49.774	47.923
13)	MA Dieldrin	5.749	6.424	662.8E6	146.4E6	50.948	47.917
14)	MA Endrin	6.006	6.640	553.6E6	114.6E6	47.428	45.875
15)	B Endosulfa...	6.279	6.851	543.7E6	119.9E6	49.148	46.218
16)	A 4,4'-DDD	6.132	6.767	530.9E6	115.0E6	53.120	50.375
17)	MA 4,4'-DDT	6.372	7.064	448.0E6	92930825	44.141	41.722
18)	B Endrin al...	6.447	6.974	439.1E6	99684335	46.964	44.870
19)	B Endosulfa...	6.660	7.197	493.6E6	109.3E6	48.584	46.067
20)	A Methoxychlor	6.917	7.523	225.0E6	51791182	40.814	40.716
21)	B Endrin ke...	7.151	7.664	547.0E6	117.6E6	49.135	47.570
22)	Mirex	7.338	8.114	370.1E6	83767991	45.153	45.560

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041420\
Data File : PL058002.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Apr 2020 01:26
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 15 01:53:09 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
Quant Title : GC Extractables
QLast Update : Sat Apr 04 06:41:11 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

