

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041421\
 Data File : PL066146.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Apr 2021 21:49
 Operator : AR\AJ
 Sample : M2024-03MS
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 FK-GF-02-071-BMS

Manual Integrations
 APPROVED

Ankita
 4/15/2021 12:08:23 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 15 01:19:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041421.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 14 15:37:20 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.446	4.155	54120355	82949064	21.263	20.621
2)	SA Decachlor...	8.171	9.353	41221225	71695116	17.282	21.953 #
Target Compounds							
2)	A alpha-BHC	3.881	4.559	181.7E6	263.1E6	44.158	45.307
3)	MA gamma-BHC...	4.159	4.852	179.5E6	250.9E6	46.074m	47.133
4)	MA Heptachlor	4.460	5.369	181.7E6	307.1E6	43.798	55.180 #
5)	MB Aldrin	4.704	5.680	123.6E6	273.1E6	31.572	54.753 #
6)	B beta-BHC	4.408	5.028	78469624	69780114	44.527	25.741 #
7)	B delta-BHC	4.609	5.252	181.0E6	237.7E6	43.885	46.232
8)	B Heptachlo...	5.143	6.074	173.0E6	285.3E6	46.684	61.769m#
9)	A Endosulfan I	5.478	6.439	160.9E6	184.0E6	44.928	43.975
10)	B gamma-Chl...	5.369	6.311	334.3E6	425.1E6	90.297	93.144m
11)	B alpha-Chl...	5.425	6.384	347.6E6	509.6E6	93.858	112.750m
12)	B 4,4'-DDE	5.592	6.541	180.4E6	224.4E6	48.349	54.204
13)	MA Dieldrin	5.718	6.701	203.2E6	189.3E6	50.152	45.311
14)	MA Endrin	5.972	6.923	168.9E6	144.8E6	48.436	44.822
15)	B Endosulfa...	6.245	7.137	177.6E6	149.9E6	54.086	43.008
16)	A 4,4'-DDD	6.103	7.040	114.9E6	156.9E6	34.897	45.846 #
17)	MA 4,4'-DDT	6.342	7.341	146.9E6	172.0E6	46.078	50.701
18)	B Endrin al...	6.411	7.263	105.6E6	114.2E6	40.949	40.817
19)	B Endosulfa...	6.623	7.487	123.3E6	157.4E6	35.856	44.553
20)	A Methoxychlor	6.888	7.801	66390350	72387287	36.669	40.397
21)	B Endrin ke...	7.108	7.967	125.9E6	140.6E6	34.234	38.224m
22)	Mirex	7.297	8.423	100.1E6	114.2E6	34.989	40.331

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041421\
 Data File : PL066146.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Apr 2021 21:49
 Operator : AR\AJ
 Sample : M2024-03MS
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 FK-GF-02-071-BMS

Manual Integrations
 APPROVED

Ankita
 4/15/2021 12:08:23 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 15 01:19:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041421.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 14 15:37:20 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

