

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041422\
 Data File : PL074067.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Apr 2022 17:12
 Operator : AR\AJ
 Sample : N2155-05
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 22L076-E

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 04/18/2022
 Supervised By : mohammad ahmed 04/18/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 18 05:45:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040822.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 14 09:29:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.638	5.491	4007497	20566977	12.403	11.535
28)	SA Decachlor...	10.412	11.335	6718245	21274998	12.919	12.045
Target Compounds							
4)	MA Heptachlor	6.183	7.100	420429	1241480	0.933	0.476m#
11)	B alpha-Chl...	7.429	8.283	713825	7040021	1.616	3.270 #
12)	B 4,4'-DDE	7.644	8.462	2332153	12351923	6.393	6.202
16)	A 4,4'-DDD	8.229	8.998	366649	4014727	1.135	2.609m#
17)	MA 4,4'-DDT	8.489	9.307	3900564	11366936	10.552	6.412m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041422\
Data File : PL074067.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2022 17:12
Operator : AR\AJ
Sample : N2155-05
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
22L076-E

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 04/18/2022
Supervised By : mohammad ahmed 04/18/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 18 05:45:16 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040822.M
Quant Title : GC Extractables
QLast Update : Thu Apr 14 09:29:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

