

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041425\
 Data File : PL095214.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Apr 2025 17:24
 Operator : AR\AJ
 Sample : PCHLORICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PCHLORICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 14 17:46:26 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
 Quant Title : GC Extractables
 QLast Update : Mon Apr 14 17:38:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.534	2.766	15794833	24163549	5.783	5.256
28)	SA Decachlor...	9.051	7.899	14297947	23998031	5.984	5.469
Target Compounds							
23)	Chlordane-1	4.695	3.762	7291356	8080201	56.084	53.837
24)	Chlordane-2	5.226	4.339	7635219	9455504	57.034	54.529
25)	Chlordane-3	5.936	4.968	29234892	26006110	59.637	51.567
26)	Chlordane-4	6.018	5.030	34422840	25674735	59.441	51.461
27)	Chlordane-5	6.868	5.926	5974550	9679336	54.565	51.807



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041425\
Data File : PL095214.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 17:24
Operator : AR\AJ
Sample : PCHLORICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 17:46:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041425.M
Quant Title : GC Extractables
QLast Update : Mon Apr 14 17:38:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.534 min
Delta R.T.: 0.000 min
Response: 15794833
Conc: 5.78 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC050




#1 Tetrachloro-m-xylene

R.T.: 2.766 min
Delta R.T.: 0.000 min
Response: 24163549
Conc: 5.26 ng/ml




#23 Chlordane-1

R.T.: 4.695 min
Delta R.T.: 0.000 min
Response: 7291356
Conc: 56.08 ng/ml



#23 Chlordane-1


R.T.: 3.762 min
Delta R.T.: 0.000 min
Response: 8080201
Conc: 53.84 ng/ml



#24 Chlordane-2


R.T.: 5.226 min
Delta R.T.: 0.000 min
Response: 7635219
Conc: 57.03 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC050




#24 Chlordane-2

R.T.: 4.339 min
Delta R.T.: 0.000 min
Response: 9455504
Conc: 54.53 ng/ml




#25 Chlordane-3

R.T.: 5.936 min
Delta R.T.: 0.000 min
Response: 29234892
Conc: 59.64 ng/ml



#25 Chlordane-3


R.T.: 4.968 min
Delta R.T.: 0.000 min
Response: 26006110
Conc: 51.57 ng/ml



#26 Chlordane-4


R.T.: 6.018 min
Delta R.T.: 0.000 min
Response: 34422840
Conc: 59.44 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC050




#26 Chlordane-4

R.T.: 5.030 min
Delta R.T.: 0.000 min
Response: 25674735
Conc: 51.46 ng/ml




#27 Chlordane-5

R.T.: 6.868 min
Delta R.T.: 0.000 min
Response: 5974550
Conc: 54.56 ng/ml



#27 Chlordane-5


R.T.: 5.926 min
Delta R.T.: 0.000 min
Response: 9679336
Conc: 51.81 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.051 min
Delta R.T.: 0.000 min
Response: 14297947
Conc: 5.98 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC050



#28 Decachlorobiphenyl

R.T.: 7.899 min
Delta R.T.: 0.000 min
Response: 23998031
Conc: 5.47 ng/ml