

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041519\
 Data File : PL047407.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Apr 2019 22:29
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 4/16/2019 11:52:22 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 16 01:51:45 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.357	3.976	362.1E6	112.4E6	49.781	49.070
28)	SA Decachlor...	8.034	9.030	405.9E6	120.6E6	48.258	51.997
Target Compounds							
2)	A alpha-BHC	3.786	4.366	601.1E6	161.3E6	51.544	50.207
3)	MA gamma-BHC...	4.063	4.650	546.3E6	149.5E6	50.843	49.966
4)	MA Heptachlor	4.350	5.155	565.4E6	155.9E6	51.281	50.676m
5)	MB Aldrin	4.590	5.460	530.4E6	143.5E6	51.418	50.085
6)	B beta-BHC	4.317	4.824	222.4E6	66832193	49.395	50.413
7)	B delta-BHC	4.512	5.039	526.3E6	131.7E6	51.995	56.767
8)	B Heptachlor...	5.027	5.846	497.1E6	135.9E6	50.962	51.524m
9)	A Endosulfan I	5.359	6.203	471.8E6	127.2E6	50.179	48.733
10)	B gamma-Chl...	5.250	6.081	524.5E6	142.5E6	51.059	50.226
11)	B alpha-Chl...	5.306	6.154	511.2E6	139.0E6	50.419	48.987
12)	B 4,4'-DDE	5.472	6.312	426.1E6	122.8E6	46.285m	49.450
13)	MA Dieldrin	5.599	6.460	506.1E6	130.7E6	50.430	48.216
14)	MA Endrin	5.852	6.678	468.1E6	118.3E6	49.693	49.505
15)	B Endosulfa...	6.127	6.888	443.7E6	117.8E6	49.517	48.830
16)	A 4,4'-DDD	5.985	6.802	388.7E6	100.6E6	51.426	49.987
17)	MA 4,4'-DDT	6.219	7.098	388.1E6	103.6E6	51.245	49.396m
18)	B Endrin al...	6.293	7.012	361.8E6	101.8E6	48.274	47.937
19)	B Endosulfa...	6.505	7.235	401.2E6	112.3E6	48.862	47.589
20)	A Methoxychlor	6.764	7.558	203.8E6	61086773	47.943	47.920
21)	B Endrin ke...	6.988	7.704	450.0E6	120.1E6	48.667	48.332
22)	Mirex	7.164	8.155	341.3E6	97601763	46.844	48.797

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041519\
Data File : PL047407.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Apr 2019 22:29
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
4/16/2019 11:52:22 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 16 01:51:45 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

