

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041522\
 Data File : PL074142.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Apr 2022 03:57
 Operator : AR\AJ
 Sample : N2434-01
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 RECOVERED-CRUSHED-AGG

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 04/18/2022
 Supervised By : mohammad ahmed 04/18/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 16 05:22:34 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040822.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 14 09:29:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.637	5.491	9329617	51183771	28.874	28.707
28)	SA Decachlor...	10.409	11.334	13404822	46341570	25.777	26.237
Target Compounds							
13)	MA Dieldrin	7.779	8.613	477555	1136159	1.112	0.567m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041522\
Data File : PL074142.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Apr 2022 03:57
Operator : AR\AJ
Sample : N2434-01
Misc :
ALS Vial : 41 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
RECOVERED-CRUSHED-AGG


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 04/18/2022
Supervised By : mohammad ahmed 04/18/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 16 05:22:34 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040822.M
Quant Title : GC Extractables
QLast Update : Thu Apr 14 09:29:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 4.637 min
Delta R.T.: -0.002 min
Response: 9329617
Conc: 28.87 ng/ml

Instrument :
ECD_L
ClientSampleId :
RECOVERED-CRUSHED-AGG


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 04/18/2022
Supervised By :mohammad ahmed 04/18/2022




#1 Tetrachloro-m-xylene

R.T.: 5.491 min
Delta R.T.: 0.000 min
Response: 51183771
Conc: 28.71 ng/ml




#13 Dieldrin

R.T.: 7.779 min
Delta R.T.: 0.006 min
Response: 477555
Conc: 1.11 ng/ml



#13 Dieldrin

R.T.: 8.613 min
Delta R.T.: 0.000 min
Response: 1136159
Conc: 0.57 ng/ml m




#28 Decachlorobiphenyl

R.T.: 10.409 min
Delta R.T.: -0.004 min
Response: 13404822
Conc: 25.78 ng/ml

Instrument :
ECD_L
ClientSampleId :
RECOVERED-CRUSHED-AGG

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 04/18/2022
Supervised By :mohammad ahmed 04/18/2022



#28 Decachlorobiphenyl

R.T.: 11.334 min
Delta R.T.: -0.002 min
Response: 46341570
Conc: 26.24 ng/ml