

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041523\
 Data File : PL082077.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Apr 2023 19:53
 Operator : AR\AJ
 Sample : 02307-01
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PL-02-041223

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 04/14/2023
 Supervised By : Ankita Jodhani 04/14/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 13 22:57:08 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 23 01:07:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.325	2.653	23997503	23300110	10.318	11.673
28) SA Decachlor...	8.746	7.765	22319640	26009126	12.472	13.114m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041523\
Data File : PL082077.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Apr 2023 19:53
Operator : AR\AJ
Sample : 02307-01
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PL-02-041223


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 04/14/2023
Supervised By : Ankita Jodhani 04/14/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 13 22:57:08 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
Quant Title : GC Extractables
QLast Update : Thu Mar 23 01:07:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.325 min
Delta R.T.: -0.010 min
Response: 23997503
Conc: 10.32 ng/ml

Instrument :
ECD_L
ClientSampleId :
PL-02-041223


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 04/14/2023
Supervised By :Ankita Jodhani 04/14/2023




#1 Tetrachloro-m-xylene

R.T.: 2.653 min
Delta R.T.: -0.003 min
Response: 23300110
Conc: 11.67 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.746 min
Delta R.T.: -0.014 min
Response: 22319640
Conc: 12.47 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.765 min
Delta R.T.: -0.014 min
Response: 26009126
Conc: 13.11 ng/ml m