

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041619\
 Data File : PL047485.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Apr 2019 04:13
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 4/17/2019 4:53:25 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 17 04:52:28 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.356	3.974	373.5E6	111.9E6	51.341	48.854
28)	SA Decachlor...	8.033	9.026	422.0E6	117.8E6	50.178	50.825
Target Compounds							
2)	A alpha-BHC	3.786	4.364	615.8E6	161.2E6	52.807	50.155
3)	MA gamma-BHC...	4.063	4.648	562.1E6	148.8E6	52.315	49.704
4)	MA Heptachlor	4.350	5.153	572.9E6	152.7E6	51.966	49.634m
5)	MB Aldrin	4.590	5.458	545.2E6	141.0E6	52.851	49.216
6)	B beta-BHC	4.316	4.820	230.6E6	69391827	51.206	52.344m
7)	B delta-BHC	4.512	5.035	545.9E6	135.0E6	53.923	58.194m
8)	B Heptachlor...	5.027	5.843	505.8E6	134.4E6	51.860	50.964m
9)	A Endosulfan I	5.358	6.200	478.0E6	125.9E6	50.837	48.225
10)	B gamma-Chl...	5.249	6.078	536.5E6	140.4E6	52.225	49.477
11)	B alpha-Chl...	5.306	6.150	520.3E6	136.8E6	51.320	48.232
12)	B 4,4'-DDE	5.472	6.308	376.0E6	120.3E6	40.835m	48.459m
13)	MA Dieldrin	5.596	6.456	515.7E6	129.5E6	51.391m	47.786m
14)	MA Endrin	5.851	6.673	463.9E6	107.8E6	49.254	45.128m
15)	B Endosulfa...	6.126	6.884	452.6E6	115.7E6	50.511	47.931
16)	A 4,4'-DDD	5.984	6.797	400.9E6	100.5E6	53.035	49.945m
17)	MA 4,4'-DDT	6.220	7.095	373.9E6	96775806	49.362	46.151m
18)	B Endrin al...	6.293	7.009	372.7E6	100.3E6	49.727	47.200
19)	B Endosulfa...	6.504	7.231	413.4E6	108.6E6	50.347	46.030
20)	A Methoxychlor	6.764	7.555	199.7E6	57140868	46.990	44.824
21)	B Endrin ke...	6.986	7.701	479.6E6	120.4E6	51.866m	48.471
22)	Mirex	7.163	8.152	351.4E6	94992510	48.235	47.492

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041619\
Data File : PL047485.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2019 04:13
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
4/17/2019 4:53:25 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 17 04:52:28 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

