

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041621\
 Data File : PL066269.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Apr 2021 23:51
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 11:32:05 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041421.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 14 15:37:20 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.445	4.155	122.5E6	204.9E6	48.136	50.927
2)	SA Decachlor...	8.170	9.356	120.8E6	159.5E6	50.630	48.834
Target Compounds							
2)	A alpha-BHC	3.880	4.560	207.0E6	309.7E6	50.312	53.336
3)	MA gamma-BHC...	4.160	4.853	197.0E6	283.3E6	50.559	53.218
4)	MA Heptachlor	4.459	5.369	210.8E6	286.6E6	50.827	51.498
5)	MB Aldrin	4.704	5.681	201.9E6	260.6E6	51.548	52.232
6)	B beta-BHC	4.408	5.029	87787768	133.8E6	49.814	53.785
7)	B delta-BHC	4.609	5.252	211.4E6	265.9E6	51.254	51.706
8)	B Heptachlo...	5.143	6.076	191.0E6	227.9E6	51.557	49.333
9)	A Endosulfan I	5.477	6.440	194.2E6	215.6E6	54.251	51.520
10)	B gamma-Chl...	5.367	6.313	203.5E6	236.2E6	54.967	51.756
11)	B alpha-Chl...	5.425	6.386	200.4E6	233.4E6	54.095	51.632
12)	B 4,4'-DDE	5.591	6.541	201.9E6	215.8E6	54.131	52.129
13)	MA Dieldrin	5.717	6.702	219.3E6	217.8E6	54.124	52.127
14)	MA Endrin	5.972	6.925	178.5E6	165.5E6	51.203	51.225
15)	B Endosulfa...	6.242	7.139	172.2E6	177.4E6	52.433	50.889
16)	A 4,4'-DDD	6.102	7.041	170.6E6	177.4E6	51.820	51.833
17)	MA 4,4'-DDT	6.342	7.342	161.6E6	174.5E6	50.700	51.426
18)	B Endrin al...	6.410	7.264	131.7E6	141.1E6	51.076	50.446
19)	B Endosulfa...	6.622	7.489	176.6E6	180.7E6	51.371	51.155
20)	A Methoxychlor	6.887	7.802	92270557	90718191	50.964	50.627
21)	B Endrin ke...	7.108	7.970	189.6E6	186.3E6	51.558	50.638
22)	Mirex	7.296	8.425	137.8E6	139.0E6	48.175	49.091

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041621\
Data File : PL066269.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Apr 2021 23:51
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 11:32:05 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041421.M
Quant Title : GC Extractables
QLast Update : Wed Apr 14 15:37:20 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

