

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041719\
 Data File : PL047496.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Apr 2019 13:04
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 4/18/2019 10:44:47 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 18 01:10:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.357	3.976	349.2E6	115.7E6	48.008	50.548
28)	SA Decachlor...	8.032	9.026	414.9E6	119.9E6	49.329	51.710
Target Compounds							
2)	A alpha-BHC	3.786	4.365	581.9E6	164.8E6	49.897	51.279
3)	MA gamma-BHC...	4.063	4.649	514.6E6	152.3E6	47.894	50.898
4)	MA Heptachlor	4.349	5.155	531.5E6	152.8E6	48.203	49.680
5)	MB Aldrin	4.589	5.459	503.3E6	144.6E6	48.791	50.479
6)	B beta-BHC	4.316	4.823	211.7E6	67148908	47.009	50.652
7)	B delta-BHC	4.511	5.038	500.1E6	121.5E6	49.401	52.353
8)	B Heptachlo...	5.026	5.843	467.9E6	127.1E6	47.976	48.199m
9)	A Endosulfan I	5.358	6.201	451.3E6	129.8E6	47.996	49.752
10)	B gamma-Chl...	5.249	6.079	491.8E6	141.4E6	47.875	49.819
11)	B alpha-Chl...	5.305	6.151	483.5E6	141.2E6	47.683	49.788
12)	B 4,4'-DDE	5.471	6.310	451.4E6	125.0E6	49.025m	50.365
13)	MA Dieldrin	5.597	6.456	485.6E6	134.8E6	48.389	49.728m
14)	MA Endrin	5.850	6.675	411.1E6	107.8E6	43.645	45.113
15)	B Endosulfa...	6.125	6.886	427.7E6	120.5E6	47.738	49.936
16)	A 4,4'-DDD	5.983	6.800	369.6E6	102.2E6	48.891	50.792
17)	MA 4,4'-DDT	6.218	7.096	354.0E6	101.8E6	46.740	48.540m
18)	B Endrin al...	6.292	7.010	365.6E6	106.8E6	48.778	50.273
19)	B Endosulfa...	6.503	7.232	391.5E6	114.6E6	47.682	48.580
20)	A Methoxychlor	6.763	7.556	191.2E6	60093448	44.988	47.140
21)	B Endrin ke...	6.985	7.702	469.6E6	127.5E6	50.789m	51.322
22)	Mirex	7.162	8.153	334.4E6	98131486	45.900	49.061

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041719\
Data File : PL047496.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2019 13:04
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
4/18/2019 10:44:47 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 18 01:10:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

