

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041823\
 Data File : PL082171.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2023 14:20
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 18 22:21:01 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 23 01:07:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.326	2.650	114.3E6	98743660	49.136	49.467
2)	SA Decachlor...	8.751	7.766	84862271	94129647	47.419	47.462m
Target Compounds							
2)	A alpha-BHC	3.776	3.148	174.4E6	143.8E6	50.868	49.319
3)	MA gamma-BHC...	4.102	3.476	162.2E6	133.2E6	51.063m	47.827
4)	MA Heptachlor	4.681	3.808	155.5E6	140.7E6	49.337m	49.094
5)	MB Aldrin	5.020	4.086	143.4E6	130.2E6	50.138	48.856
6)	B beta-BHC	4.307	3.776	72379103	59430702	48.090	49.900
7)	B delta-BHC	4.549	4.003	157.9E6	132.8E6	51.938	51.334
8)	B Heptachlo...	5.446	4.588	136.9E6	121.2E6	50.321	49.888
9)	A Endosulfan I	5.828	4.955	120.6E6	109.0E6	48.812	49.392
10)	B gamma-Chl...	5.701	4.837	130.0E6	120.9E6	49.536	49.742
11)	B alpha-Chl...	5.779	4.900	128.5E6	120.4E6	49.653	49.557
12)	B 4,4'-DDE	5.958	5.091	114.3E6	107.2E6	50.704	54.430
13)	MA Dieldrin	6.102	5.220	120.5E6	111.3E6	49.787	49.473
14)	MA Endrin	6.330	5.495	101.4E6	97458556	47.920	48.158
15)	B Endosulfa...	6.552	5.790	98051744	93223032	46.205	51.390m
16)	A 4,4'-DDD	6.474	5.646	91500959	83396517	50.736	54.955
17)	MA 4,4'-DDT	6.785	5.895	94638715	88615833	46.628	51.423
18)	B Endrin al...	6.682	5.972	80409404	77495728	49.147	49.901
19)	B Endosulfa...	6.917	6.193	101.5E6	97605545	48.705	50.165m
20)	A Methoxychlor	7.267	6.475	50288799	48721794	46.698	48.916
21)	B Endrin ke...	7.397	6.700	104.9E6	108.0E6	50.064	50.521
22)	Mirex	7.861	6.870	80995222	91551008	47.275	47.689m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041823\
Data File : PL082171.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2023 14:20
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 18 22:21:01 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
Quant Title : GC Extractables
QLast Update : Thu Mar 23 01:07:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

