

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041919\
 Data File : PL047587.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Apr 2019 17:03
 Operator : AJ\SJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL041919

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 17:14:08 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041919.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 19 16:51:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.333	3.940	462.3E6	125.5E6	52.102	47.022
28)	SA Decachlor...	8.007	8.977	500.8E6	137.1E6	49.577	45.159
Target Compounds							
2)	A alpha-BHC	3.762	4.328	748.9E6	181.0E6	53.400	48.606
3)	MA gamma-BHC...	4.038	4.612	629.0E6	166.0E6	51.977	46.972
4)	MA Heptachlor	4.324	5.115	693.1E6	166.1E6	52.001	45.075
5)	MB Aldrin	4.563	5.418	660.8E6	156.1E6	53.887	48.104
6)	B beta-BHC	4.292	4.785	281.4E6	77080805	51.362	46.464
7)	B delta-BHC	4.487	4.999	675.3E6	132.5E6	54.569	49.375
8)	B Heptachlor...	5.001	5.804	524.5E6	130.6E6	44.795	47.186
9)	A Endosulfan I	5.331	6.159	590.9E6	140.7E6	52.854	47.449
10)	B gamma-Chl...	5.223	6.038	644.7E6	151.2E6	52.709	47.839
11)	B alpha-Chl...	5.279	6.110	632.4E6	148.7E6	52.617	46.460
12)	B 4,4'-DDE	5.448	6.270	616.0E6	140.9E6	52.965	47.434
13)	MA Dieldrin	5.571	6.416	634.8E6	147.7E6	53.293	47.943
14)	MA Endrin	5.824	6.633	599.4E6	132.0E6	53.077	47.826
15)	B Endosulfa...	6.099	6.843	558.2E6	132.1E6	52.718	47.639
16)	A 4,4'-DDD	5.959	6.759	481.2E6	114.2E6	52.999	48.040
17)	MA 4,4'-DDT	6.194	7.056	485.7E6	120.2E6	56.157	48.361
18)	B Endrin al...	6.266	6.967	458.2E6	114.1E6	52.061	46.902
19)	B Endosulfa...	6.477	7.190	500.1E6	126.7E6	51.520	47.500
20)	A Methoxychlor	6.739	7.516	251.2E6	68740949	52.363	47.643
21)	B Endrin ke...	6.960	7.659	560.8E6	134.1E6	53.191	47.435
22)	Mirex	7.134	8.109	413.3E6	106.1E6	51.125	46.694

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041919\
Data File : PL047587.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Apr 2019 17:03
Operator : AJ\SJ
Sample : PSTDICV050
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL041919

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 17:14:08 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041919.M
Quant Title : GC Extractables
QLast Update : Fri Apr 19 16:51:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

