

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042020\
 Data File : PL058217.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Apr 2020 00:30
 Operator : AJ\MA
 Sample : L2327-15
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 05:10:16 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 04 06:41:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.457	3.956	237.4E6	53471298	23.554	24.640
28)	SA Decachlor...	8.205	8.978	212.6E6	44794728	20.625	20.053
Target Compounds							
4)	MA Heptachlor	4.472	5.140	937389	828908	0.058	0.247 #
7)	B delta-BHC	0.000	4.985f	0	7413309	N.D.	4.871 #
10)	B gamma-Chl...	0.000	6.043	0	422225	N.D.	0.130 #
15)	B Endosulfa...	0.000	6.869f	0	767647	N.D.	0.296 #
24)	Chlordane-2	4.847f	0.000	1845616	0	4.572	N.D. #
25)	Chlordane-3	0.000	6.043	0	422225	N.D.	0.976 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042020\
Data File : PL058217.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Apr 2020 00:30
Operator : AJ\MA
Sample : L2327-15
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 05:10:16 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
Quant Title : GC Extractables
QLast Update : Sat Apr 04 06:41:11 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.457 min
Delta R.T.: -0.001 min
Response: 237449066
Conc: 23.55 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.956 min
Delta R.T.: 0.000 min
Response: 53471298
Conc: 24.64 ng/ml




#4 Heptachlor

R.T.: 4.472 min
Delta R.T.: -0.006 min
Response: 937389
Conc: 0.06 ng/ml



#4 Heptachlor


R.T.: 5.140 min
Delta R.T.: 0.012 min
Response: 828908
Conc: 0.25 ng/ml



#7 delta-BHC


R.T.: 0.000 min
Exp R.T.: 4.634 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#7 delta-BHC

R.T.: 4.985 min
Delta R.T.: -0.029 min
Response: 7413309
Conc: 4.87 ng/ml




#10 gamma-Chlordane

R.T.: 0.000 min
Exp R.T.: 5.397 min
Response: 0
Conc: N.D.



#10 gamma-Chlordane


R.T.: 6.043 min
Delta R.T.: -0.006 min
Response: 422225
Conc: 0.13 ng/ml



#15 Endosulfan II


R.T.: 0.000 min
Exp R.T.: 6.282 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 6.869 min
Delta R.T.: 0.017 min
Response: 767647
Conc: 0.30 ng/ml




#24 Chlordane-2

R.T.: 4.847 min
Delta R.T.: 0.018 min
Response: 1845616
Conc: 4.57 ng/ml



#24 Chlordane-2


R.T.: 0.000 min
Exp R.T.: 5.407 min
Response: 0
Conc: N.D.



#25 Chloroform-3


R.T.: 0.000 min
Exp R.T.: 5.398 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#25 Chloroform-3

R.T.: 6.043 min
Delta R.T.: -0.006 min
Response: 422225
Conc: 0.98 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.205 min
Delta R.T.: -0.002 min
Response: 212610873
Conc: 20.62 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.978 min
Delta R.T.: -0.002 min
Response: 44794728
Conc: 20.05 ng/ml