

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042020\
 Data File : PL058220.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Apr 2020 01:13
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 05:10:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 04 06:41:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.457	3.956	484.8E6	103.1E6	48.090	47.510
28)	SA Decachlor...	8.204	8.979	441.1E6	89927288	42.793	40.258
Target Compounds							
2)	A alpha-BHC	3.894	4.343	847.6E6	164.0E6	48.792	49.930
3)	MA gamma-BHC...	4.177	4.626	788.8E6	163.5E6	54.266	50.322
4)	MA Heptachlor	4.477	5.128	790.8E6	166.9E6	49.333	49.721
5)	MB Aldrin	4.725	5.429	766.6E6	164.4E6	52.060	50.765
6)	B beta-BHC	4.429	4.799	323.9E6	70340268	51.295	50.656
7)	B delta-BHC	4.633	5.013	798.1E6	141.8E6	52.587	56.450
8)	B Heptachlo...	5.168	5.814	694.9E6	161.1E6	51.706	59.660
9)	A Endosulfan I	5.506	6.168	630.4E6	147.3E6	51.430	49.410
10)	B gamma-Chl...	5.395	6.048	703.9E6	161.8E6	51.830	49.941
11)	B alpha-Chl...	5.453	6.119	677.2E6	161.0E6	50.857	49.201
12)	B 4,4'-DDE	5.617	6.278	670.9E6	144.4E6	51.211	48.761
13)	MA Dieldrin	5.749	6.424	677.8E6	150.4E6	52.100	49.207
14)	MA Endrin	6.007	6.640	572.5E6	119.1E6	49.048	47.677
15)	B Endosulfa...	6.280	6.851	550.1E6	120.4E6	49.725	46.420
16)	A 4,4'-DDD	6.133	6.768	535.0E6	113.3E6	53.533	49.636
17)	MA 4,4'-DDT	6.373	7.064	473.1E6	94987384	46.613	42.645
18)	B Endrin al...	6.448	6.975	463.7E6	101.0E6	49.591	45.452
19)	B Endosulfa...	6.661	7.198	491.2E6	108.1E6	48.348	45.571
20)	A Methoxychlor	6.918	7.524	236.5E6	51360943	42.907	40.378
21)	B Endrin ke...	7.152	7.665	570.8E6	115.0E6	51.275	46.505
22)	Mirex	7.339	8.115	361.0E6	81391680	44.036	44.268

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042020\
Data File : PL058220.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Apr 2020 01:13
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 05:10:59 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
Quant Title : GC Extractables
QLast Update : Sat Apr 04 06:41:11 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

