

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042020\
 Data File : PL058198.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2020 19:46
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 05:04:35 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 04 06:41:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.457	3.956	482.8E6	105.2E6	47.889	48.495
28)	SA Decachlor...	8.205	8.980	419.0E6	89708837	40.645	40.160
Target Compounds							
2)	A alpha-BHC	3.893	4.343	841.2E6	167.2E6	48.424	50.932
3)	MA gamma-BHC...	4.177	4.626	777.7E6	165.4E6	53.504	50.913
4)	MA Heptachlor	4.477	5.127	785.3E6	168.9E6	48.991	50.337
5)	MB Aldrin	4.725	5.429	756.5E6	165.7E6	51.376	51.174
6)	B beta-BHC	4.429	4.799	319.0E6	71301088	50.515	51.348
7)	B delta-BHC	4.633	5.013	788.8E6	140.7E6	51.975	56.030
8)	B Heptachlo...	5.168	5.814	688.8E6	157.2E6	51.252	58.210
9)	A Endosulfan I	5.506	6.168	616.3E6	146.3E6	50.281	49.072
10)	B gamma-Chl...	5.395	6.048	687.5E6	161.6E6	50.620	49.876
11)	B alpha-Chl...	5.453	6.119	663.6E6	160.0E6	49.838	48.903
12)	B 4,4'-DDE	5.617	6.279	648.2E6	143.8E6	49.479	48.569
13)	MA Dieldrin	5.749	6.425	651.0E6	150.5E6	50.043	49.258
14)	MA Endrin	6.007	6.641	543.4E6	119.3E6	46.549	47.746
15)	B Endosulfa...	6.280	6.852	534.8E6	121.3E6	48.349	46.755
16)	A 4,4'-DDD	6.133	6.768	501.5E6	113.2E6	50.176	49.620
17)	MA 4,4'-DDT	6.373	7.065	467.9E6	97162924	46.105	43.622
18)	B Endrin al...	6.448	6.976	447.2E6	100.9E6	47.833	45.424
19)	B Endosulfa...	6.661	7.198	458.2E6	110.4E6	45.098	46.504
20)	A Methoxychlor	6.918	7.524	228.4E6	52725833	41.435	41.451
21)	B Endrin ke...	7.152	7.667	543.0E6	115.4E6	48.777	46.686
22)	Mirex	7.338	8.116	338.3E6	81575733	41.267	44.368



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042020\
Data File : PL058198.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Apr 2020 19:46
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 05:04:35 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
Quant Title : GC Extractables
QLast Update : Sat Apr 04 06:41:11 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.457 min
Delta R.T.: -0.001 min
Response: 482770565
Conc: 47.89 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 3.956 min
Delta R.T.: -0.002 min
Response: 105241404
Conc: 48.50 ng/ml




#2 alpha-BHC

R.T.: 3.893 min
Delta R.T.: -0.002 min
Response: 841217324
Conc: 48.42 ng/ml



#2 alpha-BHC


R.T.: 4.343 min
Delta R.T.: -0.002 min
Response: 167244520
Conc: 50.93 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.177 min
Delta R.T.: -0.002 min
Response: 777698172
Conc: 53.50 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 4.626 min
Delta R.T.: -0.001 min
Response: 165393785
Conc: 50.91 ng/ml




#4 Heptachlor

R.T.: 4.477 min
Delta R.T.: -0.002 min
Response: 785348164
Conc: 48.99 ng/ml



#4 Heptachlor


R.T.: 5.127 min
Delta R.T.: -0.002 min
Response: 168939456
Conc: 50.34 ng/ml



#5 Aldrin


R.T.: 4.725 min
Delta R.T.: -0.001 min
Response: 756503889
Conc: 51.38 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 5.429 min
Delta R.T.: -0.002 min
Response: 165710067
Conc: 51.17 ng/ml




#6 beta-BHC

R.T.: 4.429 min
Delta R.T.: -0.001 min
Response: 319016644
Conc: 50.52 ng/ml



#6 beta-BHC


R.T.: 4.799 min
Delta R.T.: -0.001 min
Response: 71301088
Conc: 51.35 ng/ml



#7 delta-BHC


R.T.: 4.633 min
Delta R.T.: -0.001 min
Response: 788788990
Conc: 51.98 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 5.013 min
Delta R.T.: -0.001 min
Response: 140722199
Conc: 56.03 ng/ml




#8 Heptachlor epoxide

R.T.: 5.168 min
Delta R.T.: -0.001 min
Response: 688816865
Conc: 51.25 ng/ml



#8 Heptachlor epoxide


R.T.: 5.814 min
Delta R.T.: -0.001 min
Response: 157187006
Conc: 58.21 ng/ml



#9 Endosulfan I


R.T.: 5.506 min
Delta R.T.: -0.002 min
Response: 616272339
Conc: 50.28 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 6.168 min
Delta R.T.: -0.002 min
Response: 146320524
Conc: 49.07 ng/ml




#10 gamma-Chlordane

R.T.: 5.395 min
Delta R.T.: -0.002 min
Response: 687477754
Conc: 50.62 ng/ml



#10 gamma-Chlordane


R.T.: 6.048 min
Delta R.T.: -0.002 min
Response: 161591571
Conc: 49.88 ng/ml



#11 alpha-Chlordane


R.T.: 5.453 min
Delta R.T.: -0.002 min
Response: 663631548
Conc: 49.84 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 6.119 min
Delta R.T.: -0.001 min
Response: 160028256
Conc: 48.90 ng/ml




#12 4,4'-DDE

R.T.: 5.617 min
Delta R.T.: -0.001 min
Response: 648175286
Conc: 49.48 ng/ml



#12 4,4'-DDE


R.T.: 6.279 min
Delta R.T.: -0.002 min
Response: 143797214
Conc: 48.57 ng/ml



#13 Dieldrin


R.T.: 5.749 min
Delta R.T.: -0.001 min
Response: 651004188
Conc: 50.04 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 6.425 min
Delta R.T.: 0.000 min
Response: 150515131
Conc: 49.26 ng/ml




#14 Endrin

R.T.: 6.007 min
Delta R.T.: -0.001 min
Response: 543361442
Conc: 46.55 ng/ml



#14 Endrin


R.T.: 6.641 min
Delta R.T.: -0.001 min
Response: 119294045
Conc: 47.75 ng/ml



#15 Endosulfan II


R.T.: 6.280 min
Delta R.T.: -0.002 min
Response: 534837247
Conc: 48.35 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 6.852 min
Delta R.T.: 0.000 min
Response: 121265142
Conc: 46.75 ng/ml




#16 4,4'-DDD

R.T.: 6.133 min
Delta R.T.: -0.001 min
Response: 501480858
Conc: 50.18 ng/ml



#16 4,4'-DDD


R.T.: 6.768 min
Delta R.T.: 0.000 min
Response: 113240437
Conc: 49.62 ng/ml



#17 4,4'-DDT


R.T.: 6.373 min
Delta R.T.: -0.001 min
Response: 467939449
Conc: 46.10 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 7.065 min
Delta R.T.: 0.000 min
Response: 97162924
Conc: 43.62 ng/ml




#18 Endrin aldehyde

R.T.: 6.448 min
Delta R.T.: -0.002 min
Response: 447217893
Conc: 47.83 ng/ml



#18 Endrin aldehyde


R.T.: 6.976 min
Delta R.T.: 0.000 min
Response: 100914698
Conc: 45.42 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.661 min
Delta R.T.: -0.002 min
Response: 458180227
Conc: 45.10 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 7.198 min
Delta R.T.: -0.001 min
Response: 110355334
Conc: 46.50 ng/ml




#20 Methoxychlor

R.T.: 6.918 min
Delta R.T.: -0.002 min
Response: 228416990
Conc: 41.44 ng/ml



#20 Methoxychlor


R.T.: 7.524 min
Delta R.T.: 0.000 min
Response: 52725833
Conc: 41.45 ng/ml



#21 Endrin ketone


R.T.: 7.152 min
Delta R.T.: -0.001 min
Response: 543018215
Conc: 48.78 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 7.667 min
Delta R.T.: 0.000 min
Response: 115407838
Conc: 46.69 ng/ml




#22 Mirex

R.T.: 7.338 min
Delta R.T.: -0.002 min
Response: 338272521
Conc: 41.27 ng/ml



#22 Mirex


R.T.: 8.116 min
Delta R.T.: 0.000 min
Response: 81575733
Conc: 44.37 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.205 min
Delta R.T.: -0.002 min
Response: 418987242
Conc: 40.64 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 8.980 min
Delta R.T.: -0.001 min
Response: 89708837
Conc: 40.16 ng/ml