

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042219\
 Data File : PL047588.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2019 08:19
 Operator : AJ\SJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: 00 Apr 23 00:59:00 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041919.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 19 16:51:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds

14) MA	Endrin	5.796f	0.000	13874033	0	1.229	N.D.	#
17) MA	4,4'-DDT	6.176	0.000	71400917	0	8.256	N.D.	#
18) B	Endrin al...	6.257	0.000	29712409	0	3.376	N.D.	#
20) A	Methoxychlor	6.747	0.000	26784182	0	5.583	N.D.	#
21) B	Endrin ke...	6.955	0.000	25358888	0	2.405	N.D.	#



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042219\
Data File : PL047588.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2019 08:19
Operator : AJ\SJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 00:59:00 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041919.M
Quant Title : GC Extractables
QLast Update : Fri Apr 19 16:51:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 3.330 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
HEXANE




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.941 min
Response: 0
Conc: N.D.




#14 Endrin

R.T.: 5.796 min
Delta R.T.: -0.025 min
Response: 13874033
Conc: 1.23 ng/ml



#14 Endrin


R.T.: 0.000 min
Exp R.T. : 6.631 min
Response: 0
Conc: N.D.



#17 4,4'-DDT


R.T.: 6.176 min
Delta R.T.: -0.014 min
Response: 71400917
Conc: 8.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
HEXANE




#17 4,4'-DDT

R.T.: 0.000 min
Exp R.T. : 7.055 min
Response: 0
Conc: N.D.




#18 Endrin aldehyde

R.T.: 6.257 min
Delta R.T.: -0.006 min
Response: 29712409
Conc: 3.38 ng/ml



#18 Endrin aldehyde


R.T.: 0.000 min
Exp R.T. : 6.966 min
Response: 0
Conc: N.D.



#20 Methoxychlor


R.T.: 6.747 min
Delta R.T.: 0.011 min
Response: 26784182
Conc: 5.58 ng/ml

Instrument :
ECD_L
ClientSampleId :
HEXANE




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T. : 7.515 min
Response: 0
Conc: N.D.




#21 Endrin ketone

R.T.: 6.955 min
Delta R.T.: 0.000 min
Response: 25358888
Conc: 2.41 ng/ml



#21 Endrin ketone


R.T.: 0.000 min
Exp R.T. : 7.658 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.003 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
HEXANE



#28 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.975 min
Response: 0
Conc: N.D.