

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042219\
 Data File : PL047591.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2019 09:20
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 00:59:35 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041919.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 19 16:51:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.331	3.940	488.4E6	129.1E6	55.047	48.350
28)	SA Decachlor...	8.003	8.973	476.3E6	127.4E6	47.154	41.969
Target Compounds							
2)	A alpha-BHC	3.760	4.327	791.6E6	181.7E6	56.450	48.805
3)	MA gamma-BHC...	4.036	4.611	722.7E6	167.5E6	59.714	47.397
4)	MA Heptachlor	4.321	5.114	731.0E6	172.6E6	54.851	46.849
5)	MB Aldrin	4.561	5.416	699.8E6	158.4E6	57.065	48.797
6)	B beta-BHC	4.290	4.784	296.3E6	78574906	54.086	47.364
7)	B delta-BHC	4.484	4.998	703.6E6	147.6E6	56.857	54.994
8)	B Heptachlo...	4.998	5.803	629.4E6	146.6E6	53.755	52.940
9)	A Endosulfan I	5.329	6.158	605.6E6	139.1E6	54.166	46.900
10)	B gamma-Chl...	5.221	6.036	670.4E6	153.5E6	54.806	48.572
11)	B alpha-Chl...	5.277	6.108	654.7E6	151.8E6	54.470	47.410
12)	B 4,4'-DDE	5.445	6.268	612.9E6	139.1E6	52.704	46.829
13)	MA Dieldrin	5.569	6.414	645.8E6	146.2E6	54.218	47.440
14)	MA Endrin	5.822	6.631	586.5E6	128.3E6	51.936	46.492
15)	B Endosulfa...	6.096	6.841	557.5E6	127.9E6	52.645	46.115
16)	A 4,4'-DDD	5.956	6.757	485.6E6	112.3E6	53.488	47.246
17)	MA 4,4'-DDT	6.191	7.054	490.5E6	114.1E6	56.720	45.907
18)	B Endrin al...	6.263	6.965	453.4E6	111.0E6	51.511	45.641
19)	B Endosulfa...	6.474	7.188	500.2E6	120.8E6	51.523	45.280
20)	A Methoxychlor	6.736	7.514	249.2E6	63437814	51.936	43.968
21)	B Endrin ke...	6.956	7.656	549.3E6	122.7E6	52.104	43.396
22)	Mirex	7.131	8.107	400.7E6	102.0E6	49.571	44.870

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042219\
Data File : PL047591.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2019 09:20
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 00:59:35 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041919.M
Quant Title : GC Extractables
QLast Update : Fri Apr 19 16:51:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

