

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042219\
 Data File : PL047606.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2019 13:05
 Operator : AJ\SJ
 Sample : K2475-01MSD
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-TRACK-8-9MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 01:03:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041919.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 19 16:51:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.331	3.940	192.1E6	60603365	21.654	22.703
28)	SA Decachlor...	8.003	8.973	146.3E6	40109943	14.481	13.213
Target Compounds							
2)	A alpha-BHC	3.759	4.327	789.2E6	187.8E6	56.279	50.452
3)	MA gamma-BHC...	4.035	4.611	735.5E6	177.2E6	60.773	50.151
4)	MA Heptachlor	4.321	5.114	738.4E6	177.0E6	55.400	48.029
5)	MB Aldrin	4.560	5.416	695.8E6	158.8E6	56.737	48.924
6)	B beta-BHC	4.289	4.784	301.9E6	81364343	55.114	49.046
7)	B delta-BHC	4.483	4.998	517.9E6	89879020	41.855	33.495
8)	B Heptachlo...	4.997	5.802	624.7E6	136.2E6	53.352	49.187
9)	A Endosulfan I	5.329	6.157	569.5E6	133.8E6	50.941	45.102
10)	B gamma-Chl...	5.220	6.035	646.4E6	159.5E6	52.845	50.468
11)	B alpha-Chl...	5.277	6.108	606.2E6	144.4E6	50.437	45.115
12)	B 4,4'-DDE	5.445	6.267	598.0E6	129.6E6	51.423	43.633
13)	MA Dieldrin	5.568	6.414	654.8E6	142.5E6	54.974	46.247
14)	MA Endrin	5.821	6.631	552.2E6	125.2E6	48.900	45.370
15)	B Endosulfa...	6.096	6.841	496.6E6	120.6E6	46.897	43.492
16)	A 4,4'-DDD	5.956	6.757	430.0E6	106.8E6	47.365	44.898
17)	MA 4,4'-DDT	6.190	7.054	430.3E6	106.6E6	49.751	42.875
18)	B Endrin al...	6.262	6.965	383.5E6	101.2E6	43.568	41.594
19)	B Endosulfa...	6.474	7.188	441.0E6	112.2E6	45.429	42.057
20)	A Methoxychlor	6.736	7.513	225.8E6	62033774	47.070	42.995
21)	B Endrin ke...	6.956	7.656	489.1E6	125.5E6	46.391	44.387
22)	Mirex	7.131	8.106	342.3E6	93315252	42.342	41.052

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042219\
Data File : PL047606.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2019 13:05
Operator : AJ\SJ
Sample : K2475-01MSD
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-TRACK-8-9MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 01:03:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041919.M
Quant Title : GC Extractables
QLast Update : Fri Apr 19 16:51:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

