

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042219\
 Data File : PL047625.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2019 18:09
 Operator : AJ\SJ
 Sample : K2515-07MSD
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-10MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 01:08:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041919.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 19 16:51:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.331	3.940	225.2E6	63031783	25.383	23.612
28)	SA Decachlor...	8.003	8.973	190.0E6	55670927	18.807	18.339
Target Compounds							
2)	A alpha-BHC	3.759	4.328	947.8E6	220.3E6	67.583	59.171
3)	MA gamma-BHC...	4.035	4.611	879.8E6	205.9E6	72.699	58.271
4)	MA Heptachlor	4.321	5.115	892.0E6	209.8E6	66.928	56.953
5)	MB Aldrin	4.560	5.417	853.3E6	196.4E6	69.586	60.500
6)	B beta-BHC	4.289	4.785	357.4E6	95159882	65.244	57.362
7)	B delta-BHC	4.483	4.999	704.5E6	140.5E6	56.927	52.371
8)	B Heptachlo...	4.997	5.803	744.0E6	173.0E6	63.544	62.508
9)	A Endosulfan I	5.328	6.158	683.6E6	161.8E6	61.149	54.542
10)	B gamma-Chl...	5.220	6.037	779.4E6	179.9E6	63.722	56.903
11)	B alpha-Chl...	5.276	6.109	756.9E6	179.1E6	62.974	55.943
12)	B 4,4'-DDE	5.444	6.268	722.4E6	167.9E6	62.120	56.522
13)	MA Dieldrin	5.567	6.414	720.4E6	174.3E6	60.480	56.576
14)	MA Endrin	5.820	6.631	658.0E6	148.9E6	58.268	53.952
15)	B Endosulfa...	6.095	6.841	623.8E6	142.2E6	58.912	51.273
16)	A 4,4'-DDD	5.956	6.757	557.2E6	129.7E6	61.366	54.556
17)	MA 4,4'-DDT	6.190	7.054	590.9E6	132.1E6	68.323	53.150
18)	B Endrin al...	6.262	6.965	480.9E6	115.7E6	54.639	47.575
19)	B Endosulfa...	6.474	7.188	526.9E6	130.2E6	54.282	48.807
20)	A Methoxychlor	6.735	7.514	266.9E6	73533972	55.634	50.965
21)	B Endrin ke...	6.955	7.656	584.6E6	147.5E6	55.446	52.139
22)	Mirex	7.130	8.106	436.4E6	117.8E6	53.991	51.810

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042219\
Data File : PL047625.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2019 18:09
Operator : AJ\SJ
Sample : K2515-07MSD
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-10MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 01:08:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041919.M
Quant Title : GC Extractables
QLast Update : Fri Apr 19 16:51:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

