

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX042221.M
 Title : GC Extractables
 Last Update : Fri Apr 23 00:49:23 2021
 Response Via : Initial Calibration

Calibration Files

500 =PL066433.D 1000 =PL066431.D 250 =PL066434.D
 750 =PL066432.D 100 =PL066435.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.794	2.756	2.743	2.798	2.841	2.787	E6	1.39
2)	Toxaphene-1	1.750	1.589	1.764	1.691	1.903	1.739	E4	6.60
3)	Toxaphene-2	1.855	1.807	1.879	1.822	2.019	1.876	E4	4.50
4)	Toxaphene-3	5.006	4.783	4.894	5.003	5.075	4.952	E4	2.31
5)	Toxaphene-4	3.284	3.210	3.131	3.360	3.107	3.218	E4	3.28
6)	Toxaphene-5	5.259	5.165	5.109	5.256	5.118	5.181	E4	1.40
7) SA	Decachlorobiphenyl	2.467	2.329	2.517	2.414	2.599	2.465	E6	4.15

Signal #2 Calibration Files

500 =PL066433.D 1000 =PL066431.D 250 =PL066434.D
 750 =PL066432.D 100 =PL066435.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	4.302	4.315	4.150	4.303	4.287	4.272	E6	1.61
2)	Toxaphene-1	3.493	3.318	3.251	3.491	3.440	3.398	E4	3.20
3)	Toxaphene-2	3.593	3.483	3.460	3.584	3.663	3.556	E4	2.36
4)	Toxaphene-3	6.595	6.397	6.283	6.657	6.616	6.510	E4	2.49
5)	Toxaphene-4	6.359	6.344	6.093	6.248	6.313	6.271	E4	1.73
6)	Toxaphene-5	6.507	6.486	6.217	6.625	6.745	6.516	E4	3.02
7) SA	Decachlorobiphenyl	3.463	3.298	3.494	3.384	3.805	3.489	E6	5.52

(#) = Out of Range