

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042222\
 Data File : PL074294.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2022 11:07
 Operator : AR\AJ
 Sample : PB143979BL (Sig #1); PB143979BS (Sig #2)
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB143979BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 23:04:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040822.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 14 09:29:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.642	5.486	6995762	36750240	21.651	20.612
28)	SA Decachlor...	10.407	11.329	10869699	39135611	20.902	22.157

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042222\
Data File : PL074294.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2022 11:07
Operator : AR\AJ
Sample : PB143979BL (Sig #1); PB143979BS (Sig #2)
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB143979BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 23:04:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040822.M
Quant Title : GC Extractables
QLast Update : Thu Apr 14 09:29:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.642 min
Delta R.T.: 0.003 min
Response: 6995762
Conc: 21.65 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB143979BL




#1 Tetrachloro-m-xylene

R.T.: 5.486 min
Delta R.T.: -0.006 min
Response: 36750240
Conc: 20.61 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.407 min
Delta R.T.: -0.007 min
Response: 10869699
Conc: 20.90 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.329 min
Delta R.T.: -0.007 min
Response: 39135611
Conc: 22.16 ng/ml