

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042319\
 Data File : PL047669.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Apr 2019 20:43
 Operator : AJ\SJ
 Sample : K2532-23
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-4

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 24 01:03:13 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041919.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 19 16:51:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.330	3.942	156.7E6	40440901	17.657	15.150
28)	SA Decachlor...	8.003	8.981	115.2E6	32151854	11.408	10.592
Target Compounds							
6)	B beta-BHC	0.000	4.770f	0	1111809	N.D.	0.670 #
15)	B Endosulfa...	0.000	6.865f	0	905255	N.D.	0.326 #
19)	B Endosulfa...	6.456f	7.195	5953491	2113942	0.613	0.792 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042319\
Data File : PL047669.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Apr 2019 20:43
Operator : AJ\SJ
Sample : K2532-23
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-4

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 24 01:03:13 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041919.M
Quant Title : GC Extractables
QLast Update : Fri Apr 19 16:51:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.330 min
Delta R.T.: 0.000 min
Response: 156662870
Conc: 17.66 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-4




#1 Tetrachloro-m-xylene

R.T.: 3.942 min
Delta R.T.: 0.000 min
Response: 40440901
Conc: 15.15 ng/ml




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.288 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 4.770 min
Delta R.T.: -0.015 min
Response: 1111809
Conc: 0.67 ng/ml



#15 Endosulfan II


R.T.: 0.000 min
Exp R.T.: 6.095 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
TP-4




#15 Endosulfan II

R.T.: 6.865 min
Delta R.T.: 0.022 min
Response: 905255
Conc: 0.33 ng/ml




#19 Endosulfan Sulfate

R.T.: 6.456 min
Delta R.T.: -0.018 min
Response: 5953491
Conc: 0.61 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.195 min
Delta R.T.: 0.006 min
Response: 2113942
Conc: 0.79 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.003 min
Delta R.T.: 0.000 min
Response: 115240855
Conc: 11.41 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-4



#28 Decachlorobiphenyl

R.T.: 8.981 min
Delta R.T.: 0.007 min
Response: 32151854
Conc: 10.59 ng/ml