

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042319\
 Data File : PL047682.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Apr 2019 00:00
 Operator : AJ\SJ
 Sample : K2532-08
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 24 01:06:55 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041919.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 19 16:51:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.330	3.942	195.3E6	50052949	22.012	18.750
28) SA Decachlor...	8.003	8.982	219.0E6	58669633	21.681	19.327
Target Compounds						
7) B delta-BHC	0.000	4.977f	0	3322681	N.D.	1.238 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042319\
Data File : PL047682.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2019 00:00
Operator : AJ\SJ
Sample : K2532-08
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 24 01:06:55 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041919.M
Quant Title : GC Extractables
QLast Update : Fri Apr 19 16:51:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.330 min
Delta R.T.: 0.000 min
Response: 195308969
Conc: 22.01 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-2




#1 Tetrachloro-m-xylene

R.T.: 3.942 min
Delta R.T.: 0.001 min
Response: 50052949
Conc: 18.75 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.483 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 4.977 min
Delta R.T.: -0.022 min
Response: 3322681
Conc: 1.24 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.003 min
Delta R.T.: 0.000 min
Response: 219005971
Conc: 21.68 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-2



#28 Decachlorobiphenyl

R.T.: 8.982 min
Delta R.T.: 0.007 min
Response: 58669633
Conc: 19.33 ng/ml