

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042319\
 Data File : PL047677.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Apr 2019 22:44
 Operator : AJ\SJ
 Sample : K2204-04MSD
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TR-05-042219MSD

Manual Integrations
 APPROVED

Sohil
 4/24/2019 11:27:29 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 24 01:05:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041919.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 19 16:51:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.330	3.942	196.5E6	55279207	22.150	20.708
28)	SA Decachlor...	8.003	8.982	221.2E6	60577886	21.895	19.956
Target Compounds							
2)	A alpha-BHC	3.759	4.329	854.3E6	202.2E6	60.919	54.304m
3)	MA gamma-BHC...	4.035	4.614	798.4E6	188.2E6	65.972	53.274
4)	MA Heptachlor	4.320	5.118	835.5E6	198.5E6	62.688	53.872
5)	MB Aldrin	4.559	5.421	808.4E6	183.6E6	65.924	56.553
6)	B beta-BHC	4.288	4.788	328.8E6	89472100	60.015	53.933
7)	B delta-BHC	4.483	5.002	678.7E6	146.9E6	54.847	54.742
8)	B Heptachlor...	4.997	5.807	745.0E6	172.5E6	63.627	62.293
9)	A Endosulfan I	5.328	6.163	702.7E6	164.0E6	62.859	55.298
10)	B gamma-Chl...	5.219	6.041	769.0E6	178.9E6	62.871	56.588
11)	B alpha-Chl...	5.275	6.113	749.0E6	175.8E6	62.316	54.912
12)	B 4,4'-DDE	5.443	6.273	725.2E6	163.3E6	62.356	54.977
13)	MA Dieldrin	5.567	6.419	761.1E6	173.2E6	63.891	56.225
14)	MA Endrin	5.820	6.636	690.6E6	151.8E6	61.154	55.014
15)	B Endosulfa...	6.095	6.845	642.7E6	152.8E6	60.693	55.115m
16)	A 4,4'-DDD	5.955	6.763	582.3E6	135.5E6	64.140	56.977
17)	MA 4,4'-DDT	6.189	7.060	549.4E6	133.0E6	63.532	53.503
18)	B Endrin al...	6.262	6.971	525.6E6	129.8E6	59.715	53.365
19)	B Endosulfa...	6.473	7.194	604.7E6	146.9E6	62.287	55.088
20)	A Methoxychlor	6.735	7.520	281.9E6	74497589	58.763	51.633
21)	B Endrin ke...	6.955	7.663	679.4E6	162.2E6	64.440	57.359
22)	Mirex	7.131	8.114	505.1E6	128.1E6	62.483	56.341

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042319\
Data File : PL047677.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Apr 2019 22:44
Operator : AJ\SJ
Sample : K2204-04MSD
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TR-05-042219MSD

Manual Integrations
APPROVED

Sohil
4/24/2019 11:27:29 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 24 01:05:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041919.M
Quant Title : GC Extractables
QLast Update : Fri Apr 19 16:51:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

