

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042420\
 Data File : PL058356.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Apr 2020 12:31
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 25 00:51:07 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 04 06:41:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.459	3.955	491.4E6	102.6E6	48.742	47.269
2)	SA Decachlor...	8.206	8.979	482.3E6	93892335	46.784	42.033
Target Compounds							
2)	A alpha-BHC	3.896	4.343	853.6E6	162.2E6	49.136	49.407
3)	MA gamma-BHC...	4.180	4.625	799.0E6	161.8E6	54.968	49.791
4)	MA Heptachlor	4.480	5.127	793.1E6	162.6E6	49.473	48.437
5)	MB Aldrin	4.728	5.428	780.0E6	160.7E6	52.972	49.622
6)	B beta-BHC	4.432	4.799	322.3E6	71748830	51.040	51.670
7)	B delta-BHC	4.636	5.012	800.0E6	146.9E6	52.712	58.394
8)	B Heptachlo...	5.170	5.814	716.5E6	155.7E6	53.312	57.673
9)	A Endosulfan I	5.509	6.168	650.4E6	144.6E6	53.067	48.502
10)	B gamma-Chl...	5.398	6.047	723.4E6	158.6E6	53.268	48.942
11)	B alpha-Chl...	5.456	6.118	708.5E6	158.2E6	53.208	48.351
12)	B 4,4'-DDE	5.619	6.277	696.6E6	141.8E6	53.179	47.906
13)	MA Dieldrin	5.752	6.424	702.3E6	147.9E6	53.984	48.406
14)	MA Endrin	6.009	6.640	563.1E6	109.2E6	48.238	43.695
15)	B Endosulfa...	6.282	6.850	559.4E6	118.7E6	50.569	45.784
16)	A 4,4'-DDD	6.136	6.767	551.9E6	111.6E6	55.224	48.919
17)	MA 4,4'-DDT	6.375	7.063	496.2E6	97608451	48.894	43.822
18)	B Endrin al...	6.450	6.974	486.0E6	100.8E6	51.979	45.352
19)	B Endosulfa...	6.664	7.197	491.1E6	108.5E6	48.339	45.716
20)	A Methoxychlor	6.920	7.523	255.0E6	52881689	46.261	41.573
21)	B Endrin ke...	7.154	7.665	594.8E6	119.4E6	53.429	48.313
22)	Mirex	7.340	8.114	359.5E6	85275958	43.857	46.380

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042420\
Data File : PL058356.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2020 12:31
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 25 00:51:07 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
Quant Title : GC Extractables
QLast Update : Sat Apr 04 06:41:11 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

