

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042519\
 Data File : PL047831.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Apr 2019 06:38
 Operator : AJ\SJ
 Sample : K2568-17
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 26 09:56:01 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042419.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 25 01:53:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.329	3.942	225.1E6	49665300	24.345	20.804
28)	SA Decachlor...	8.003	8.982	138.6E6	38424963	12.377	13.461
Target Compounds							
10)	B gamma-Chl...	0.000	6.030	0	5141748	N.D.	1.699 #
12)	B 4,4'-DDE	5.443	6.273	10001991	2575082	0.804	0.933
13)	MA Dieldrin	5.556	0.000	22148878	0	1.706	N.D. #
15)	B Endosulfa...	0.000	6.866f	0	1582929	N.D.	0.603 #
17)	MA 4,4'-DDT	0.000	7.071	0	1114363	N.D.	0.468 #
19)	B Endosulfa...	6.499f	0.000	5232284	0	0.469	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042519\
Data File : PL047831.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Apr 2019 06:38
Operator : AJ\SJ
Sample : K2568-17
Misc :
ALS Vial : 60 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 26 09:56:01 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042419.M
Quant Title : GC Extractables
QLast Update : Thu Apr 25 01:53:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.329 min
Delta R.T.: 0.000 min
Response: 225149776
Conc: 24.35 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.942 min
Delta R.T.: 0.000 min
Response: 49665300
Conc: 20.80 ng/ml




#10 gamma-Chlordane

R.T.: 0.000 min
Exp R.T. : 5.218 min
Response: 0
Conc: N.D.



#10 gamma-Chlordane


R.T.: 6.030 min
Delta R.T.: -0.011 min
Response: 5141748
Conc: 1.70 ng/ml



#12 4,4'-DDE


R.T.: 5.443 min
Delta R.T.: 0.000 min
Response: 10001991
Conc: 0.80 ng/ml

Instrument :
ECD_L
ClientSampleId :




#12 4,4'-DDE

R.T.: 6.273 min
Delta R.T.: 0.000 min
Response: 2575082
Conc: 0.93 ng/ml




#13 Dieldrin

R.T.: 5.556 min
Delta R.T.: -0.010 min
Response: 22148878
Conc: 1.71 ng/ml



#13 Dieldrin


R.T.: 0.000 min
Exp R.T.: 6.420 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 0.000 min
Exp R.T.: 6.093 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 6.866 min
Delta R.T.: 0.018 min
Response: 1582929
Conc: 0.60 ng/ml




#17 4,4'-DDT

R.T.: 0.000 min
Exp R.T.: 6.188 min
Response: 0
Conc: N.D.



#17 4,4'-DDT


R.T.: 7.071 min
Delta R.T.: 0.011 min
Response: 1114363
Conc: 0.47 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.499 min
Delta R.T.: 0.026 min
Response: 5232284
Conc: 0.47 ng/ml

Instrument :
ECD_L
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 0.000 min
Exp R.T. : 7.194 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.003 min
Delta R.T.: 0.001 min
Response: 138551378
Conc: 12.38 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.982 min
Delta R.T.: 0.000 min
Response: 38424963
Conc: 13.46 ng/ml