

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042519\
 Data File : PL047778.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Apr 2019 16:03
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 4/26/2019 12:47:09 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 26 01:05:26 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042419.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 25 01:53:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.328	3.941	510.2E6	129.1E6	55.169	54.077
28)	SA Decachlor...	8.001	8.980	488.1E6	136.1E6	43.600	47.673
Target Compounds							
2)	A alpha-BHC	3.756	4.330	828.7E6	185.7E6	56.686	55.963
3)	MA gamma-BHC...	4.033	4.613	764.8E6	173.3E6	57.697	55.252
4)	MA Heptachlor	4.318	5.117	786.5E6	178.3E6	55.199	53.060
5)	MB Aldrin	4.557	5.420	777.7E6	169.4E6	55.849	54.264
6)	B beta-BHC	4.286	4.787	310.6E6	77684152	54.967	53.916
7)	B delta-BHC	4.480	5.001	740.7E6	150.6E6	57.017	59.937
8)	B Heptachlo...	4.994	5.806	693.3E6	158.2E6	54.104	56.594
9)	A Endosulfan I	5.325	6.162	656.3E6	147.4E6	54.267	52.498
10)	B gamma-Chl...	5.217	6.041	726.5E6	161.2E6	54.570	53.275
11)	B alpha-Chl...	5.273	6.113	704.6E6	158.8E6	53.998	52.499
12)	B 4,4'-DDE	5.441	6.272	654.5E6	146.3E6	52.616	52.972
13)	MA Dieldrin	5.564	6.419	708.3E6	153.6E6	54.560	52.525
14)	MA Endrin	5.817	6.635	627.0E6	129.3E6	52.758	50.603
15)	B Endosulfa...	6.092	6.846	595.6E6	133.4E6	53.194	50.836
16)	A 4,4'-DDD	5.952	6.762	547.5E6	119.1E6	55.677	53.060
17)	MA 4,4'-DDT	6.186	7.059	543.7E6	121.5E6	54.536	51.020
18)	B Endrin al...	6.259	6.970	493.9E6	116.5E6	52.345	50.416
19)	B Endosulfa...	6.470	7.193	576.8E6	135.7E6	51.744	49.565
20)	A Methoxychlor	6.733	7.520	266.0E6	67141894	50.018	47.885
21)	B Endrin ke...	6.952	7.662	630.9E6	141.1E6	53.198m	50.180
22)	Mirex	7.128	8.112	440.2E6	110.4E6	46.901	47.459

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042519\
Data File : PL047778.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Apr 2019 16:03
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

mohammad
4/26/2019 12:47:09 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 26 01:05:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042419.M
Quant Title : GC Extractables
QLast Update : Thu Apr 25 01:53:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

