

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042623\
 Data File : PL082346.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Apr 2023 15:32
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 26 21:37:50 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042023.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 21 03:12:31 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.322	2.647	134.8E6	117.3E6	50.468	50.624
28)	SA Decachlor...	8.741	7.758	94357629	104.3E6	49.426	45.693
Target Compounds							
2)	A alpha-BHC	3.772	3.145	205.7E6	174.2E6	51.677	51.182
3)	MA gamma-BHC...	4.100	3.471	191.0E6	163.1E6	51.095	50.723
4)	MA Heptachlor	4.678	3.804	185.3E6	167.2E6	50.721	48.832
5)	MB Aldrin	5.015	4.080	172.3E6	163.2E6	52.257	51.814
6)	B beta-BHC	4.303	3.771	84811760	69438556	49.728	50.893
7)	B delta-BHC	4.545	3.998	186.3E6	159.0E6	51.795	51.821
8)	B Heptachlo...	5.441	4.582	154.0E6	144.5E6	50.499	50.175
9)	A Endosulfan I	5.822	4.949	140.2E6	132.9E6	50.570	49.796
10)	B gamma-Chl...	5.696	4.830	157.7E6	146.0E6	52.456	50.298
11)	B alpha-Chl...	5.773	4.893	151.9E6	145.0E6	50.899	49.957
12)	B 4,4'-DDE	5.952	5.084	132.4E6	124.2E6	52.084	54.014
13)	MA Dieldrin	6.096	5.213	141.9E6	135.0E6	51.413	50.008
14)	MA Endrin	6.323	5.487	117.1E6	114.6E6	48.307	46.872
15)	B Endosulfa...	6.546	5.784	116.4E6	111.7E6	47.808	49.575
16)	A 4,4'-DDD	6.468	5.638	112.4E6	104.1E6	55.470	55.665
17)	MA 4,4'-DDT	6.778	5.887	104.9E6	98576351	45.854	46.661
18)	B Endrin al...	6.675	5.964	94564087	90862335	50.189	48.554
19)	B Endosulfa...	6.910	6.187	118.0E6	115.0E6	49.744	48.864
20)	A Methoxychlor	7.259	6.466	55564004	54823459	45.532	45.240
21)	B Endrin ke...	7.389	6.691	125.8E6	127.3E6	52.536	50.158
22)	Mirex	7.854	6.862	92901282	104.1E6	47.488	45.170

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042623\
Data File : PL082346.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Apr 2023 15:32
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 26 21:37:50 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042023.M
Quant Title : GC Extractables
QLast Update : Fri Apr 21 03:12:31 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

