

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042719\
 Data File : PL047914.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Apr 2019 06:03
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 4/29/2019 4:32:54 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 06:31:52 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042419.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 25 01:53:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.330	3.944	491.2E6	130.9E6	53.111	54.817
28)	SA Decachlor...	8.005	8.987	459.8E6	127.8E6	41.072	44.776
Target Compounds							
2)	A alpha-BHC	3.759	4.333	790.0E6	187.7E6	54.039	56.555
3)	MA gamma-BHC...	4.035	4.616	724.8E6	173.7E6	54.680	55.382
4)	MA Heptachlor	4.321	5.121	730.9E6	174.2E6	51.295	51.819
5)	MB Aldrin	4.560	5.424	726.5E6	166.8E6	52.173	53.432
6)	B beta-BHC	4.289	4.791	294.7E6	78122927	52.160	54.221
7)	B delta-BHC	4.484	5.005	698.0E6	152.0E6	53.728	60.506
8)	B Heptachlor...	4.997	5.811	639.2E6	152.4E6	49.877	54.538
9)	A Endosulfan I	5.329	6.167	599.7E6	141.8E6	49.589	50.486
10)	B gamma-Chl...	5.220	6.045	662.3E6	155.8E6	49.743	51.479
11)	B alpha-Chl...	5.276	6.117	643.6E6	153.3E6	49.324	50.668
12)	B 4,4'-DDE	5.445	6.277	633.1E6	141.7E6	50.900	51.324
13)	MA Dieldrin	5.568	6.423	642.4E6	147.3E6	49.488	50.373
14)	MA Endrin	5.822	6.640	540.8E6	119.3E6	45.510	46.707
15)	B Endosulfa...	6.096	6.851	524.6E6	126.8E6	46.858	48.341
16)	A 4,4'-DDD	5.956	6.767	501.2E6	118.6E6	50.973	52.821
17)	MA 4,4'-DDT	6.191	7.064	435.7E6	104.4E6	43.706	43.839
18)	B Endrin al...	6.263	6.975	432.2E6	110.9E6	45.803	48.011
19)	B Endosulfa...	6.475	7.198	515.4E6	128.1E6	46.235	46.779
20)	A Methoxychlor	6.736	7.524	227.7E6	60811715	42.815	43.370
21)	B Endrin ke...	6.956	7.667	570.2E6	136.1E6	48.082m	48.400
22)	Mirex	7.132	8.118	398.8E6	105.2E6	42.483	45.224

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042719\
 Data File : PL047914.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Apr 2019 06:03
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 4/29/2019 4:32:54 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 06:31:52 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042419.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 25 01:53:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

