

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042825\
 Data File : PL095424.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Apr 2025 10:51
 Operator : AR\AJ
 Sample : INDA201
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 28 12:31:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042825CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Apr 28 12:25:16 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.592	2.903	33718170	34044760	9.946	10.203
27)	SA Decachlor...	9.115	8.081	63352453	70475773	20.184	20.192
Target Compounds							
2)	A alpha-BHC	4.042	3.412	49839586	46727106	9.634	9.803
3)	MA gamma-BHC...	4.373	3.745	47977709	45239527	9.734	9.831
4)	MA Heptachlor	4.971	4.099	47255923	45495606	9.782	9.977
9)	A Endosulfan I	6.117	5.260	39455112	38744043	9.906	10.036
13)	MA Dieldrin	6.390	5.525	84200086	84706179	19.505	19.684
14)	MA Endrin	6.617	5.800	63787087	71316333	19.199	19.800
16)	A 4,4'-DDD	6.747	5.942	63125552	68835924	19.475	19.591
17)	MA 4,4'-DDT	7.062	6.196	56366711	67239334	19.150	19.378
20)	A Methoxychlor	7.533	6.767	140.0E6	184.0E6	99.017	102.999

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042825\
Data File : PL095424.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Apr 2025 10:51
Operator : AR\AJ
Sample : INDA201
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 28 12:31:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042825CLP.M
Quant Title : GC Extractables
QLast Update : Mon Apr 28 12:25:16 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

